

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071524\
 Data File : PL090531.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jul 2024 14:14
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 15 14:33:05 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071524.M
 Quant Title : GC Extractables
 QLast Update : Mon Jul 15 14:16:48 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.549	2.788	101.5E6	86692134	48.750	51.512
2)	SA Decachlor...	9.060	7.932	63614517	72928231	47.872	50.016
Target Compounds							
2)	A alpha-BHC	4.004	3.292	143.8E6	127.9E6	48.455	52.923
3)	MA gamma-BHC...	4.336	3.623	135.3E6	119.6E6	48.298	51.870
4)	MA Heptachlor	4.925	3.963	121.5E6	114.0E6	47.748	51.691
5)	MB Aldrin	5.267	4.244	119.5E6	111.4E6	47.811	52.345
6)	B beta-BHC	4.532	3.921	60177163	51606986	47.439	49.997
7)	B delta-BHC	4.779	4.151	125.3E6	115.0E6	46.728	52.297
8)	B Heptachlo...	5.692	4.746	109.6E6	100.2E6	46.611	51.384
9)	A Endosulfan I	6.077	5.116	101.5E6	93360872	47.454	51.052
10)	B gamma-Chl...	5.948	4.996	110.9E6	104.3E6	47.701	51.428
11)	B alpha-Chl...	6.027	5.060	108.8E6	103.7E6	47.400	51.315
12)	B 4,4'-DDE	6.200	5.249	96185392	92425073	47.569	51.717
13)	MA Dieldrin	6.353	5.381	106.2E6	101.0E6	47.788	51.491
14)	MA Endrin	6.581	5.656	87644117	87913135	47.673	51.079
15)	B Endosulfa...	6.801	5.950	91760582	86756114	46.653	50.782
16)	A 4,4'-DDD	6.716	5.804	80604049	76945212	47.745	51.356
17)	MA 4,4'-DDT	7.030	6.055	80389608	75962814	47.757	51.846
18)	B Endrin al...	6.930	6.129	70741808	66275678	47.682	50.726
19)	B Endosulfa...	7.164	6.352	84139483	82186013	47.909	50.968
20)	A Methoxychlor	7.505	6.629	41640563	41048577	47.163	50.768
21)	B Endrin ke...	7.648	6.858	92758574	95499011	48.590	52.257
22)	Mirex	8.122	7.040	69636820	75740202	49.170	50.949

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071524\
Data File : PL090531.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jul 2024 14:14
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 15 14:33:05 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071524.M
Quant Title : GC Extractables
QLast Update : Mon Jul 15 14:16:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

