

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071525\
 Data File : PL096363.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jul 2025 11:49
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 16 01:36:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL070725.M
 Quant Title : GC Extractables
 QLast Update : Mon Jul 07 15:22:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.535	2.825	195.8E6	261.9E6	53.670	45.387
28)	SA Decachlor...	9.020	7.994	143.5E6	210.4E6	50.996	42.200
Target Compounds							
2)	A alpha-BHC	3.983	3.331	308.3E6	403.0E6	59.946	47.780
3)	MA gamma-BHC...	4.311	3.663	294.4E6	371.9E6	59.127	47.219
4)	MA Heptachlor	4.904	4.013	264.2E6	351.5E6	57.660	46.204
5)	MB Aldrin	5.244	4.296	273.0E6	342.5E6	56.474	46.951
6)	B beta-BHC	4.497	3.960	120.1E6	158.9E6	57.781	46.398
7)	B delta-BHC	4.744	4.193	266.3E6	362.5E6	59.417	47.067
8)	B Heptachlo...	5.664	4.799	243.1E6	309.7E6	58.583	46.322
9)	A Endosulfan I	6.046	5.170	226.1E6	289.2E6	55.992	40.305 #
10)	B gamma-Chl...	5.918	5.051	248.6E6	327.2E6	56.190	46.614
11)	B alpha-Chl...	5.999	5.115	245.5E6	320.5E6	56.066	42.624
12)	B 4,4'-DDE	6.169	5.304	204.1E6	300.8E6	55.047	46.544
13)	MA Dieldrin	6.319	5.435	234.7E6	315.8E6	55.504	46.308
14)	MA Endrin	6.545	5.709	162.3E6	253.7E6	52.191	41.610
15)	B Endosulfa...	6.757	6.001	202.8E6	273.8E6	58.435	46.685
16)	A 4,4'-DDD	6.677	5.857	174.9E6	266.4E6	60.066	51.938
17)	MA 4,4'-DDT	6.992	6.110	155.0E6	242.6E6	51.052	42.304
18)	B Endrin al...	6.886	6.180	145.3E6	206.1E6	64.974	46.273 #
19)	B Endosulfa...	7.120	6.403	171.6E6	258.3E6	56.114	45.927
20)	A Methoxychlor	7.465	6.682	82812528	123.4E6	52.566	40.628
21)	B Endrin ke...	7.599	6.908	187.5E6	302.1E6	56.797	48.326
22)	Mirex	8.079	7.098	147.0E6	220.5E6	52.839	44.385
24)	Chlordane-2	5.244f	0.000	273.0E6	0	1394.328	N.D. #
25)	Chlordane-3	5.918	5.051	248.6E6	327.2E6	337.468	406.712
26)	Chlordane-4	5.999	5.115	245.5E6	320.5E6	273.627	426.580 #
27)	Chlordane-5	0.000	6.001	0	273.8E6	N.D.	965.789 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071525\
Data File : PL096363.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jul 2025 11:49
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 16 01:36:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL070725.M
Quant Title : GC Extractables
QLast Update : Mon Jul 07 15:22:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.535 min
Delta R.T.: -0.002 min
Response: 195838820
Conc: 53.67 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.003 min
Response: 261937474
Conc: 45.39 ng/ml




#2 alpha-BHC

R.T.: 3.983 min
Delta R.T.: -0.001 min
Response: 308309311
Conc: 59.95 ng/ml



#2 alpha-BHC


R.T.: 3.331 min
Delta R.T.: -0.003 min
Response: 402996243
Conc: 47.78 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.311 min
Delta R.T.: -0.002 min
Response: 294430211
Conc: 59.13 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 3.663 min
Delta R.T.: -0.003 min
Response: 371892104
Conc: 47.22 ng/ml




#4 Heptachlor

R.T.: 4.904 min
Delta R.T.: -0.001 min
Response: 264152436
Conc: 57.66 ng/ml



#4 Heptachlor


R.T.: 4.013 min
Delta R.T.: -0.003 min
Response: 351506658
Conc: 46.20 ng/ml



#5 Aldrin


R.T.: 5.244 min
Delta R.T.: -0.001 min
Response: 272981719
Conc: 56.47 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 4.296 min
Delta R.T.: -0.003 min
Response: 342539320
Conc: 46.95 ng/ml




#6 beta-BHC

R.T.: 4.497 min
Delta R.T.: -0.002 min
Response: 120115486
Conc: 57.78 ng/ml



#6 beta-BHC


R.T.: 3.960 min
Delta R.T.: -0.003 min
Response: 158862400
Conc: 46.40 ng/ml



#7 delta-BHC


R.T.: 4.744 min
Delta R.T.: -0.001 min
Response: 266268481
Conc: 59.42 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 4.193 min
Delta R.T.: -0.003 min
Response: 362505475
Conc: 47.07 ng/ml




#8 Heptachlor epoxide

R.T.: 5.664 min
Delta R.T.: -0.001 min
Response: 243071667
Conc: 58.58 ng/ml



#8 Heptachlor epoxide


R.T.: 4.799 min
Delta R.T.: -0.002 min
Response: 309657477
Conc: 46.32 ng/ml



#9 Endosulfan I


R.T.: 6.046 min
Delta R.T.: 0.000 min
Response: 226053667
Conc: 55.99 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 5.170 min
Delta R.T.: -0.002 min
Response: 289157444
Conc: 40.31 ng/ml




#10 gamma-Chlordane

R.T.: 5.918 min
Delta R.T.: 0.000 min
Response: 248630882
Conc: 56.19 ng/ml



#10 gamma-Chlordane


R.T.: 5.051 min
Delta R.T.: -0.003 min
Response: 327215247
Conc: 46.61 ng/ml



#11 alpha-Chlordane


R.T.: 5.999 min
Delta R.T.: 0.000 min
Response: 245505544
Conc: 56.07 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 5.115 min
Delta R.T.: -0.004 min
Response: 320543769
Conc: 42.62 ng/ml




#12 4,4'-DDE

R.T.: 6.169 min
Delta R.T.: 0.000 min
Response: 204085057
Conc: 55.05 ng/ml



#12 4,4'-DDE


R.T.: 5.304 min
Delta R.T.: -0.001 min
Response: 300843299
Conc: 46.54 ng/ml



#13 Dieldrin


R.T.: 6.319 min
Delta R.T.: 0.000 min
Response: 234693301
Conc: 55.50 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 5.435 min
Delta R.T.: -0.001 min
Response: 315784869
Conc: 46.31 ng/ml




#14 Endrin

R.T.: 6.545 min
Delta R.T.: 0.000 min
Response: 162324280
Conc: 52.19 ng/ml



#14 Endrin


R.T.: 5.709 min
Delta R.T.: -0.002 min
Response: 253672895
Conc: 41.61 ng/ml



#15 Endosulfan II


R.T.: 6.757 min
Delta R.T.: 0.000 min
Response: 202771178
Conc: 58.43 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 6.001 min
Delta R.T.: -0.002 min
Response: 273753299
Conc: 46.68 ng/ml




#16 4,4'-DDD

R.T.: 6.677 min
Delta R.T.: 0.000 min
Response: 174894848
Conc: 60.07 ng/ml



#16 4,4'-DDD


R.T.: 5.857 min
Delta R.T.: -0.002 min
Response: 266375582
Conc: 51.94 ng/ml



#17 4,4'-DDT


R.T.: 6.992 min
Delta R.T.: 0.000 min
Response: 154954360
Conc: 51.05 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 6.110 min
Delta R.T.: -0.003 min
Response: 242619772
Conc: 42.30 ng/ml




#18 Endrin aldehyde

R.T.: 6.886 min
Delta R.T.: 0.000 min
Response: 145253102
Conc: 64.97 ng/ml



#18 Endrin aldehyde


R.T.: 6.180 min
Delta R.T.: -0.001 min
Response: 206129450
Conc: 46.27 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.120 min
Delta R.T.: 0.000 min
Response: 171643377
Conc: 56.11 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 6.403 min
Delta R.T.: -0.003 min
Response: 258309608
Conc: 45.93 ng/ml




#20 Methoxychlor

R.T.: 7.465 min
Delta R.T.: 0.000 min
Response: 82812528
Conc: 52.57 ng/ml



#20 Methoxychlor


R.T.: 6.682 min
Delta R.T.: -0.003 min
Response: 123420076
Conc: 40.63 ng/ml



#21 Endrin ketone


R.T.: 7.599 min
Delta R.T.: 0.000 min
Response: 187523666
Conc: 56.80 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 6.908 min
Delta R.T.: -0.001 min
Response: 302148236
Conc: 48.33 ng/ml




#22 Mirex

R.T.: 8.079 min
Delta R.T.: 0.000 min
Response: 147010905
Conc: 52.84 ng/ml



#22 Mirex


R.T.: 7.098 min
Delta R.T.: -0.003 min
Response: 220512083
Conc: 44.38 ng/ml



#24 Chlordane-2


R.T.: 5.244 min
Delta R.T.: 0.027 min
Response: 272981719
Conc: 1394.33 ng/m

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#24 Chlordane-2

R.T.: 0.000 min
Exp R.T. : 4.420 min
Response: 0
Conc: N.D.




#25 Chlordane-3

R.T.: 5.918 min
Delta R.T.: -0.002 min
Response: 248630882
Conc: 337.47 ng/ml



#25 Chlordane-3


R.T.: 5.051 min
Delta R.T.: -0.004 min
Response: 327215247
Conc: 406.71 ng/ml



#26 Chlordane-4


R.T.: 5.999 min
Delta R.T.: -0.006 min
Response: 245505544
Conc: 273.63 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#26 Chlordane-4

R.T.: 5.115 min
Delta R.T.: -0.003 min
Response: 320543769
Conc: 426.58 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.841 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 6.001 min
Delta R.T.: -0.011 min
Response: 273753299
Conc: 965.79 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.020 min
Delta R.T.: 0.000 min
Response: 143454985
Conc: 51.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 7.994 min
Delta R.T.: -0.003 min
Response: 210432164
Conc: 42.20 ng/ml