

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071618\
 Data File : PL038118.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jul 2018 16:37
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 7/17/2018 2:18:03 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 17 01:44:16 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
 Quant Title : GC Extractables
 QLast Update : Sat Jul 14 00:12:12 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.329	3.930	377.5E6	144.9E6	50.701	55.191
28)	SA Decachlor...	8.003	8.939	346.0E6	116.4E6	45.618	47.619
Target Compounds							
2)	A alpha-BHC	3.759	4.309	612.3E6	218.2E6	51.820	56.384
3)	MA gamma-BHC...	4.035	4.588	553.5E6	195.1E6	52.793	54.025m
4)	MA Heptachlor	4.319	5.100	549.5E6	181.9E6	51.657	56.907
5)	MB Aldrin	4.558	5.402	554.1E6	171.8E6	52.937	55.682
6)	B beta-BHC	4.289	4.751	246.5E6	86757350	52.226	57.194
7)	B delta-BHC	4.483	4.965	582.9E6	188.7E6	53.704	56.901
8)	B Heptachlo...	4.996	5.782	505.0E6	160.7E6	53.560	56.564
9)	A Endosulfan I	5.327	6.135	460.0E6	136.3E6	53.184	53.001
10)	B gamma-Chl...	5.219	6.015	525.9E6	154.5E6	53.208	53.288
11)	B alpha-Chl...	5.275	6.088	500.5E6	152.1E6	52.876	53.074
12)	B 4,4'-DDE	5.444	6.244	505.3E6	139.7E6	52.702	52.232
13)	MA Dieldrin	5.567	6.389	505.1E6	146.4E6	52.561	53.795
14)	MA Endrin	5.820	6.601	452.0E6	124.4E6	52.027	52.097m
15)	B Endosulfa...	6.095	6.806	423.5E6	122.7E6	53.402	51.092
16)	A 4,4'-DDD	5.955	6.725	405.7E6	111.5E6	55.310	52.114m
17)	MA 4,4'-DDT	6.190	7.025	351.8E6	103.2E6	47.358	48.398
18)	B Endrin al...	6.263	6.929	323.2E6	102.5E6	50.752	52.545
19)	B Endosulfa...	6.473	7.151	382.0E6	114.4E6	50.580m	52.493
20)	A Methoxychlor	6.739	7.480	183.5E6	55641493	47.703	48.866
21)	B Endrin ke...	6.956	7.615	431.5E6	122.9E6	54.999m	51.573

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071618\
Data File : PL038118.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jul 2018 16:37
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
7/17/2018 2:18:03 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 17 01:44:16 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
Quant Title : GC Extractables
QLast Update : Sat Jul 14 00:12:12 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

