

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071623\
 Data File : PL083950.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jul 2023 09:58
 Operator : AR\AJ
 Sample : 03501-05
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 A46W2

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 07/19/2023
 Supervised By : Ankita Jodhani 07/19/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 15 13:52:01 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071523CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Jul 15 05:26:27 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.470	2.689	846.8E6	6329552	408.308	6.165 #
27)	SA Decachlor...	9.026	7.870	25595635	15016788	19.473m	13.568 #
Target Compounds							
3)	MA gamma-BHC...	4.274	3.506	1501.9E6	522.9E6	532.350m	378.639m#
8)	B Heptachlo...	5.625	4.652	261.5E6	85368598	94.515m	68.199 #
10)	B trans-Chl...	5.914	4.889	83353642	69303914	31.873	56.032m#
11)	B cis-Chlor...	5.973	4.950	35745377	28301715	13.463	22.093m#
14)	MA Endrin	6.526	5.579	111.1E6	11945074	57.306m	10.488m#
17)	MA 4,4'-DDT	6.952	5.969	83853597	50525860	54.293	48.504
19)	B Endosulfa...	7.097	6.272	10151231	12315606	5.787	11.163m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071623\
Data File : PL083950.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jul 2023 09:58
Operator : AR\AJ
Sample : 03501-05
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 15 13:52:01 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071523CLP.M
Quant Title : GC Extractables
QLast Update : Sat Jul 15 05:26:27 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Reviewed By: Abdul Mirza 07/19/2023
Supervised By: Ankita Jodhani 07/19/2023

Manual Integrations
APPROVED

Instrument :
ECD_L
ClientSampled :
A46W2