

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071719\
 Data File : PL050539.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jul 2019 10:33
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 18 10:46:14 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
 Quant Title : GC Extractables
 QLast Update : Thu Jul 18 09:14:47 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.318	3.930	527.5E6	145.1E6	52.503	50.852
28)	SA Decachlor...	7.980	8.961	480.5E6	139.4E6	47.458	49.280
Target Compounds							
2)	A alpha-BHC	3.745	4.319	886.1E6	211.7E6	52.876	51.304
3)	MA gamma-BHC...	4.021	4.602	827.7E6	196.7E6	53.811	50.781
4)	MA Heptachlor	4.304	5.103	869.2E6	207.2E6	52.974	49.967
5)	MB Aldrin	4.542	5.405	790.4E6	185.8E6	51.021	49.297
6)	B beta-BHC	4.275	4.778	346.8E6	90133831	52.897	52.135
7)	B delta-BHC	4.468	4.991	857.2E6	195.6E6	52.239	50.573
8)	B Heptachlor...	4.978	5.792	735.9E6	174.7E6	49.812	45.685
9)	A Endosulfan I	5.309	6.147	697.9E6	160.4E6	47.986	46.637
10)	B gamma-Chl...	5.201	6.026	765.8E6	173.5E6	48.517	46.940
11)	B alpha-Chl...	5.256	6.097	737.5E6	171.6E6	48.344	46.810
12)	B 4,4'-DDE	5.424	6.258	722.9E6	172.4E6	48.237	47.857
13)	MA Dieldrin	5.547	6.403	721.6E6	166.5E6	46.562	47.011
14)	MA Endrin	5.800	6.620	651.1E6	142.5E6	44.228	45.933
15)	B Endosulfa...	6.075	6.833	656.4E6	147.0E6	46.902	46.237
16)	A 4,4'-DDD	5.934	6.748	558.7E6	126.5E6	46.618	47.496
17)	MA 4,4'-DDT	6.168	7.045	594.2E6	127.0E6	47.977	47.048
18)	B Endrin al...	6.242	6.957	519.1E6	119.1E6	45.000	43.814
19)	B Endosulfa...	6.453	7.180	560.3E6	134.4E6	43.981	46.529
20)	A Methoxychlor	6.713	7.506	275.3E6	74329636	44.118	50.659
21)	B Endrin ke...	6.935	7.648	594.9E6	142.6E6	43.492	49.314
22)	Mirex	7.108	8.096	401.3E6	110.2E6	40.931	47.958
24)	Chlordane-2	0.000	5.405f	0	185.8E6	N.D.	1581.651 #
25)	Chlordane-3	5.201	6.026	765.8E6	173.5E6	547.056	360.629 #
26)	Chlordane-4	5.256	6.097	737.5E6	171.6E6	559.176	295.337 #
27)	Chlordane-5	6.075	0.000	656.4E6	0	1292.259	N.D. #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071719\
Data File : PL050539.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jul 2019 10:33
Operator : SM\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 18 10:46:14 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
Quant Title : GC Extractables
QLast Update : Thu Jul 18 09:14:47 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

