

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX071719.M
 Title : GC Extractables
 Last Update : Thu Jul 18 03:03:34 2019
 Response Via : Initial Calibration

Calibration Files

500 =PL050484.D 1000 =PL050482.D 250 =PL050485.D
 750 =PL050483.D 100 =PL050486.D

Compound		500	1000	250	750	100	Avg	%RSD	

1) SA	Tetrachloro-m-xylene	1.030	1.015	1.025	1.020	1.060	1.030	E7	1.70
2)	Toxaphene-1	7.353	7.351	7.376	7.080	7.950	7.422	E4	4.30
3)	Toxaphene-2	1.962	1.960	1.881	2.019	1.821	1.929	E5	4.04
4)	Toxaphene-3	8.520	9.673	8.453	9.374	6.736	8.551	E4	13.39
5)	Toxaphene-4	7.787	7.957	7.297	7.856	6.943	7.568	E4	5.71
6)	Toxaphene-5	2.449	2.533	2.382	2.495	2.291	2.430	E5	3.95
7) SA	Decachlorobiphenyl	1.080	1.040	1.106	1.054	1.167	1.089	E7	4.60

Signal #2 Calibration Files

500 =PL050484.D 1000 =PL050482.D 250 =PL050485.D
 750 =PL050483.D 100 =PL050486.D

Compound		500	1000	250	750	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.914	2.919	2.892	2.904	2.939	2.914	E6	0.61
2)	Toxaphene-1	1.707	1.542	1.633	1.601	1.669	1.631	E4	3.89
3)	Toxaphene-2	1.894	1.912	1.849	1.802	1.840	1.859	E4	2.36
4)	Toxaphene-3	7.244	7.037	7.282	7.274	6.711	7.110	E4	3.44
5)	Toxaphene-4	4.564	4.540	4.406	4.507	4.349	4.473	E4	2.05
6)	Toxaphene-5	2.351	2.372	2.173	2.358	2.315	2.314	E4	3.53
7) SA	Decachlorobiphenyl	2.864	2.745	3.004	2.775	3.261	2.930	E6	7.19

(#) = Out of Range