

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071818\
 Data File : PL038250.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jul 2018 02:06
 Operator : AJ\SJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 19 03:57:26 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
 Quant Title : GC Extractables
 QLast Update : Sat Jul 14 00:12:12 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.329	3.930	148.2E6	59008794	19.902	22.480
28)	SA Decachlor...	8.000	8.938	149.2E6	54094769	19.674	22.130
Target Compounds							
2)	A alpha-BHC	3.758	4.309	117.5E6	40340793	9.946	10.424
3)	MA gamma-BHC...	4.034	4.588	109.9E6	38441997	10.479	10.643
4)	MA Heptachlor	4.288	0.000	54240110	0	5.099	N.D. #
6)	B beta-BHC	4.288	4.751	54240110	17513840	11.491	11.546
14)	MA Endrin	5.818	6.602	407.4E6	116.4E6	46.893	48.758
15)	B Endosulfa...	6.034f	6.826	23804848	2108304	3.002	0.878 #
16)	A 4,4'-DDD	5.954	6.724	53602375	14506897	7.307	6.778
17)	MA 4,4'-DDT	6.188	7.024	720.7E6	215.2E6	97.020	100.939
18)	B Endrin al...	0.000	6.927	0	2217509	N.D.	1.136 #
20)	A Methoxychlor	6.737	7.479	866.6E6	273.9E6	225.245	240.538
21)	B Endrin ke...	6.955	7.614	22891790	6981988	2.918	2.929
27)	Chlordane-5	5.818	6.927	407.4E6	2217509	2544.304	24.755 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071818\
Data File : PL038250.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jul 2018 02:06
Operator : AJ\SJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 19 03:57:26 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
Quant Title : GC Extractables
QLast Update : Sat Jul 14 00:12:12 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.329 min
Delta R.T.: 0.000 min
Response: 148196572
Conc: 19.90 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.930 min
Delta R.T.: 0.000 min
Response: 59008794
Conc: 22.48 ng/ml




#2 alpha-BHC

R.T.: 3.758 min
Delta R.T.: 0.000 min
Response: 117520433
Conc: 9.95 ng/ml



#2 alpha-BHC


R.T.: 4.309 min
Delta R.T.: 0.000 min
Response: 40340793
Conc: 10.42 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.034 min
Delta R.T.: 0.000 min
Response: 109864658
Conc: 10.48 ng/ml

Instrument :
ECD_L
ClientSampleId :




#3 gamma-BHC (Lindane)

R.T.: 4.588 min
Delta R.T.: 0.000 min
Response: 38441997
Conc: 10.64 ng/ml




#4 Heptachlor

R.T.: 4.288 min
Delta R.T.: -0.031 min
Response: 54240110
Conc: 5.10 ng/ml



#4 Heptachlor


R.T.: 0.000 min
Exp R.T.: 5.099 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 4.288 min
Delta R.T.: -0.001 min
Response: 54240110
Conc: 11.49 ng/ml

Instrument :
ECD_L
ClientSampleId :




#6 beta-BHC

R.T.: 4.751 min
Delta R.T.: 0.000 min
Response: 17513840
Conc: 11.55 ng/ml




#14 Endrin

R.T.: 5.818 min
Delta R.T.: -0.003 min
Response: 407427512
Conc: 46.89 ng/ml



#14 Endrin


R.T.: 6.602 min
Delta R.T.: 0.000 min
Response: 116441715
Conc: 48.76 ng/ml



#15 Endosulfan II


R.T.: 6.034 min
Delta R.T.: -0.062 min
Response: 23804848
Conc: 3.00 ng/ml

Instrument :
ECD_L
ClientSampleId :




#15 Endosulfan II

R.T.: 6.826 min
Delta R.T.: 0.020 min
Response: 2108304
Conc: 0.88 ng/ml




#16 4,4'-DDD

R.T.: 5.954 min
Delta R.T.: -0.003 min
Response: 53602375
Conc: 7.31 ng/ml



#16 4,4'-DDD


R.T.: 6.724 min
Delta R.T.: 0.000 min
Response: 14506897
Conc: 6.78 ng/ml



#17 4,4'-DDT


R.T.: 6.188 min
Delta R.T.: -0.003 min
Response: 720728001
Conc: 97.02 ng/ml

Instrument :
ECD_L
ClientSampleId :




#17 4,4'-DDT

R.T.: 7.024 min
Delta R.T.: 0.000 min
Response: 215205740
Conc: 100.94 ng/ml




#18 Endrin aldehyde

R.T.: 0.000 min
Exp R.T. : 6.264 min
Response: 0
Conc: N.D.



#18 Endrin aldehyde


R.T.: 6.927 min
Delta R.T.: 0.000 min
Response: 2217509
Conc: 1.14 ng/ml



#20 Methoxychlor


R.T.: 6.737 min
Delta R.T.: -0.003 min
Response: 866625204
Conc: 225.24 ng/ml

Instrument :
ECD_L
ClientSampleId :




#20 Methoxychlor

R.T.: 7.479 min
Delta R.T.: 0.001 min
Response: 273892795
Conc: 240.54 ng/ml




#21 Endrin ketone

R.T.: 6.955 min
Delta R.T.: -0.003 min
Response: 22891790
Conc: 2.92 ng/ml



#21 Endrin ketone


R.T.: 7.614 min
Delta R.T.: 0.000 min
Response: 6981988
Conc: 2.93 ng/ml



#27 Chlordane-5


R.T.: 5.818 min
Delta R.T.: -0.065 min
Response: 407427512
Conc: 2544.30 ng/ml

Instrument :
ECD_L
ClientSampleId :




#27 Chlordane-5

R.T.: 6.927 min
Delta R.T.: 0.046 min
Response: 2217509
Conc: 24.75 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.000 min
Delta R.T.: -0.004 min
Response: 149248407
Conc: 19.67 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.938 min
Delta R.T.: 0.000 min
Response: 54094769
Conc: 22.13 ng/ml