

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_L\Data\PL071819\  
 Data File : PL050552.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 18 Jul 2019 14:38  
 Operator : SM\AJ  
 Sample : HEXANE  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 ECD\_L  
 ClientSampleId :  
 HEXANE

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Jul 19 00:53:48 2019  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_L\methods\PL071719.M  
 Quant Title : GC Extractables  
 QLast Update : Thu Jul 18 09:14:47 2019  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

| Compound | RT#1 | RT#2 | Resp#1 | Resp#2 | ng/ml | ng/ml |
|----------|------|------|--------|--------|-------|-------|
|----------|------|------|--------|--------|-------|-------|

#### System Monitoring Compounds

##### Target Compounds

|      |              |       |        |   |         |      |         |
|------|--------------|-------|--------|---|---------|------|---------|
| 6) B | beta-BHC     | 0.000 | 4.800f | 0 | 7082435 | N.D. | 4.097 # |
| 8) B | Heptachlo... | 0.000 | 5.807  | 0 | 4216975 | N.D. | 1.103 # |


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_L\Data\PL071819\  
Data File : PL050552.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 18 Jul 2019 14:38  
Operator : SM\AJ  
Sample : HEXANE  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
ECD\_L  
ClientSampleId :  
HEXANE

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jul 19 00:53:48 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_L\methods\PL071719.M  
Quant Title : GC Extractables  
QLast Update : Thu Jul 18 09:14:47 2019  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min  
Exp R.T. : 3.318 min  
Response: 0  
Conc: N.D.

Instrument :  
ECD\_L  
ClientSampleId :  
HEXANE




#1 Tetrachloro-m-xylene

R.T.: 0.000 min  
Exp R.T. : 3.931 min  
Response: 0  
Conc: N.D.




#6 beta-BHC

R.T.: 0.000 min  
Exp R.T. : 4.276 min  
Response: 0  
Conc: N.D.



#6 beta-BHC


R.T.: 4.800 min  
Delta R.T.: 0.022 min  
Response: 7082435  
Conc: 4.10 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min  
Exp R.T. : 4.979 min  
Response: 0  
Conc: N.D.

Instrument :  
ECD\_L  
ClientSampleId :  
HEXANE




#8 Heptachlor epoxide

R.T.: 5.807 min  
Delta R.T.: 0.015 min  
Response: 4216975  
Conc: 1.10 ng/ml



#28 Decachlorobiphenyl

R.T.: 0.000 min  
Exp R.T. : 7.981 min  
Response: 0  
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 0.000 min  
Exp R.T. : 8.962 min  
Response: 0  
Conc: N.D.