

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071823\
 Data File : PL084016.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jul 2023 00:21
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 18 03:49:48 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071223.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 14 00:03:54 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.476	2.689	97405403	50420690	48.706	45.400
28)	SA Decachlor...	9.022	7.870	62441994	53886256	47.971	48.497
Target Compounds							
2)	A alpha-BHC	3.934	3.194	144.4E6	75690401	49.290	47.075
3)	MA gamma-BHC...	4.268	3.524	136.8E6	72917583	48.432	47.950
4)	MA Heptachlor	4.859	3.866	137.6E6	62516731	50.629	53.049
5)	MB Aldrin	5.202	4.146	142.3E6	70233853	49.692	46.579
6)	B beta-BHC	4.466	3.826	60756877	32110238	50.637	48.648
7)	B delta-BHC	4.713	4.054	139.8E6	73936222	52.449	48.879
8)	B Heptachlo...	5.631	4.652	126.5E6	62219631	47.384	48.451
9)	A Endosulfan I	6.017	5.022	109.7E6	58581240	46.337	47.762
10)	B gamma-Chl...	5.888	4.905	121.5E6	64565754	47.069	48.467
11)	B alpha-Chl...	5.968	4.969	118.5E6	64495975	46.597	47.876
12)	B 4,4'-DDE	6.146	5.167	99730319	60528301	49.568	49.276
13)	MA Dieldrin	6.294	5.289	111.0E6	64326435	46.461	49.393
14)	MA Endrin	6.523	5.564	88399792	56004072	46.418	50.948
15)	B Endosulfa...	6.744	5.862	85434869	56249201	45.582	48.940
16)	A 4,4'-DDD	6.666	5.725	73599284	50792519	50.118	52.945
17)	MA 4,4'-DDT	6.981	5.977	71870766	49424854	45.066	54.266
18)	B Endrin al...	6.874	6.044	69011091	45277876	47.104	50.368
19)	B Endosulfa...	7.110	6.268	80364029	54384085	46.488	49.484
20)	A Methoxychlor	7.461	6.562	36476617	27677361	46.632	57.752
21)	B Endrin ke...	7.594	6.773	81443595	66091068	46.474	51.151
22)	Mirex	8.069	6.954	64331639	50447432	44.957	49.426

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071823\
Data File : PL084016.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jul 2023 00:21
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 18 03:49:28 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071223.M
Quant Title : GC Extractables
QLast Update : Fri Jul 14 00:03:54 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

