

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071919\
 Data File : PL050653.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jul 2019 21:07
 Operator : SM\AJ
 Sample : K3615-02MS
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 HR-06-071719-AMS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 19 23:38:38 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 19 01:19:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.319	3.930	172.3E6	54147819	17.145	18.978
28)	SA Decachlor...	7.981	8.962	191.6E6	53503134	18.919	18.920
Target Compounds							
2)	A alpha-BHC	3.747	4.319	919.3E6	248.6E6	54.860	60.241
3)	MA gamma-BHC...	4.022	4.602	882.8E6	232.6E6	57.392	60.048
4)	MA Heptachlor	4.306	5.104	928.8E6	243.7E6	56.607	58.773
5)	MB Aldrin	4.544	5.406	879.7E6	224.4E6	56.787	59.557
6)	B beta-BHC	4.276	4.778	367.6E6	105.2E6	56.066	60.870
7)	B delta-BHC	4.470	4.991	865.4E6	209.1E6	52.743	54.058
8)	B Heptachlor...	4.980	5.793	836.1E6	216.3E6	56.592	56.542
9)	A Endosulfan I	5.310	6.148	822.1E6	203.4E6	56.525	59.160
10)	B gamma-Chl...	5.202	6.027	890.4E6	221.6E6	56.415	59.959
11)	B alpha-Chl...	5.258	6.098	866.0E6	217.2E6	56.764	59.265
12)	B 4,4'-DDE	5.426	6.259	852.4E6	216.9E6	56.880	60.182
13)	MA Dieldrin	5.549	6.404	884.0E6	213.9E6	57.042	60.404
14)	MA Endrin	5.802	6.621	823.4E6	184.9E6	55.933	59.619
15)	B Endosulfa...	6.077	6.834	793.8E6	187.9E6	56.718	59.091
16)	A 4,4'-DDD	5.936	6.750	701.1E6	165.8E6	58.508	62.280
17)	MA 4,4'-DDT	6.171	7.046	695.4E6	162.2E6	56.147	60.099
18)	B Endrin al...	6.244	6.958	630.5E6	153.6E6	54.650	56.503
19)	B Endosulfa...	6.455	7.181	696.4E6	167.7E6	54.665	58.078
20)	A Methoxychlor	6.715	7.507	336.4E6	89025536	53.918	60.674
21)	B Endrin ke...	6.937	7.649	769.4E6	184.7E6	56.253	63.884
22)	Mirex	7.110	8.097	500.1E6	131.6E6	51.004	57.310

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071919\
Data File : PL050653.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jul 2019 21:07
Operator : SM\AJ
Sample : K3615-02MS
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
HR-06-071719-AMS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 19 23:38:38 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
Quant Title : GC Extractables
QLast Update : Fri Jul 19 01:19:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

