

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072018\
 Data File : PL038385.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jul 2018 20:51
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 7/21/2018 11:45:40 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 21 02:10:31 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
 Quant Title : GC Extractables
 QLast Update : Sat Jul 14 00:12:12 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.335	3.928	361.4E6	141.6E6	48.531	53.944
28)	SA Decachlor...	8.006	8.939	355.8E6	124.3E6	46.907	50.844
Target Compounds							
2)	A alpha-BHC	3.764	4.307	554.0E6	214.5E6	46.891m	55.419
3)	MA gamma-BHC...	4.041	4.587	489.5E6	188.5E6	46.688m	52.194
4)	MA Heptachlor	4.326	5.098	485.4E6	179.3E6	45.635	56.099
5)	MB Aldrin	4.565	5.401	497.2E6	171.3E6	47.505	55.507
6)	B beta-BHC	4.296	4.750	221.0E6	85427910	46.822	56.318
7)	B delta-BHC	4.491	4.964	524.5E6	174.6E6	48.319	52.633
8)	B Heptachlo...	5.003	5.780	434.1E6	148.2E6	46.043	52.152
9)	A Endosulfan I	5.333	6.133	421.1E6	134.4E6	48.683m	52.273
10)	B gamma-Chl...	5.226	6.013	489.4E6	152.2E6	49.515	52.500
11)	B alpha-Chl...	5.282	6.087	454.2E6	152.3E6	47.982	53.137
12)	B 4,4'-DDE	5.451	6.243	456.8E6	139.1E6	47.642	52.035
13)	MA Dieldrin	5.574	6.387	462.4E6	144.5E6	48.119	53.085
14)	MA Endrin	5.827	6.602	420.2E6	116.6E6	48.358	48.845
15)	B Endosulfa...	6.102	6.806	422.0E6	124.9E6	53.219	52.000
16)	A 4,4'-DDD	5.962	6.725	388.0E6	112.7E6	52.889	52.637
17)	MA 4,4'-DDT	6.197	7.024	340.5E6	101.8E6	45.840	47.764
18)	B Endrin al...	6.269	6.928	313.2E6	104.5E6	49.191	53.532
19)	B Endosulfa...	6.481	7.150	373.3E6	117.3E6	49.419	53.815
20)	A Methoxychlor	6.744	7.479	177.0E6	59029978	46.002	51.841
21)	B Endrin ke...	6.961	7.613	439.4E6	125.6E6	56.010m	52.678m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072018\
Data File : PL038385.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jul 2018 20:51
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
7/21/2018 11:45:40 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 21 02:10:31 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
Quant Title : GC Extractables
QLast Update : Sat Jul 14 00:12:12 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

