

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072020\
 Data File : PL060019.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jul 2020 10:29
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 20 16:42:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071020.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 10 18:52:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.465	4.012	40356243	57071667	20.864	21.535
28)	SA Decachlor...	8.227	9.104	34738974	42845252	21.117	20.486
Target Compounds							
4)	MA Heptachlor	0.000	5.203	0	20834108	N.D.	6.193 #
5)	MB Aldrin	0.000	5.478f	0	27701292	N.D.	8.894 #
15)	B Endosulfa...	6.304	0.000	1420893	0	0.851	N.D. #
17)	MA 4,4'-DDT	0.000	7.124f	0	4571917	N.D.	2.322 #
18)	B Endrin al...	6.469	0.000	2295192	0	1.609	N.D. #
24)	Chlordane-2	0.000	5.478	0	27701292	N.D.	241.774 #
27)	Chlordane-5	6.304	0.000	1420893	0	25.042	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072020\
Data File : PL060019.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jul 2020 10:29
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 20 16:42:42 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071020.M
Quant Title : GC Extractables
QLast Update : Fri Jul 10 18:52:06 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.465 min
Delta R.T.: -0.001 min
Response: 40356243
Conc: 20.86 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.012 min
Delta R.T.: 0.000 min
Response: 57071667
Conc: 21.53 ng/ml




#4 Heptachlor

R.T.: 0.000 min
Exp R.T. : 4.490 min
Response: 0
Conc: N.D.



#4 Heptachlor


R.T.: 5.203 min
Delta R.T.: 0.005 min
Response: 20834108
Conc: 6.19 ng/ml



#5 Aldrin


R.T.: 0.000 min
Exp R.T.: 4.738 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#5 Aldrin

R.T.: 5.478 min
Delta R.T.: -0.027 min
Response: 27701292
Conc: 8.89 ng/ml




#15 Endosulfan II

R.T.: 6.304 min
Delta R.T.: 0.009 min
Response: 1420893
Conc: 0.85 ng/ml



#15 Endosulfan II


R.T.: 0.000 min
Exp R.T.: 6.936 min
Response: 0
Conc: N.D.



#17 4,4'-DDT


R.T.: 0.000 min
Exp R.T.: 6.392 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#17 4,4'-DDT

R.T.: 7.124 min
Delta R.T.: -0.025 min
Response: 4571917
Conc: 2.32 ng/ml




#18 Endrin aldehyde

R.T.: 6.469 min
Delta R.T.: 0.005 min
Response: 2295192
Conc: 1.61 ng/ml



#18 Endrin aldehyde


R.T.: 0.000 min
Exp R.T.: 7.060 min
Response: 0
Conc: N.D.



#24 Chlordane-2


R.T.: 0.000 min
Exp R.T.: 4.840 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#24 Chlordane-2

R.T.: 5.478 min
Delta R.T.: -0.001 min
Response: 27701292
Conc: 241.77 ng/ml




#27 Chlordane-5

R.T.: 6.304 min
Delta R.T.: -0.001 min
Response: 1420893
Conc: 25.04 ng/ml



#27 Chlordane-5


R.T.: 0.000 min
Exp R.T.: 7.010 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.227 min
Delta R.T.: -0.004 min
Response: 34738974
Conc: 21.12 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 9.104 min
Delta R.T.: 0.001 min
Response: 42845252
Conc: 20.49 ng/ml