

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072021\
 Data File : PL068291.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jul 2021 19:48
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 21 05:48:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071921.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 20 11:44:21 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.644	5.571	9035502	22638276	21.053	22.185
28)	SA Decachlor...	10.423	11.493	12693057	20669387	22.569	23.172
Target Compounds							
4)	MA Heptachlor	6.164f	0.000	417539	0	0.685	N.D. #
6)	B beta-BHC	6.164	0.000	417539	0	1.427	N.D. #
8)	B Heptachlo...	0.000	8.039	0	4009163	N.D.	3.546 #
10)	B gamma-Chl...	0.000	8.297	0	2866371	N.D.	2.545 #
13)	MA Dieldrin	7.782	0.000	59765	0	0.111	N.D. #
14)	MA Endrin	8.092f	0.000	536267	0	1.072	N.D. #
15)	B Endosulfa...	0.000	9.220f	0	646282	N.D.	0.713 #
25)	Chlordane-3	0.000	8.297	0	2866371	N.D.	18.508 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072021\
Data File : PL068291.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jul 2021 19:48
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 21 05:48:43 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071921.M
Quant Title : GC Extractables
QLast Update : Tue Jul 20 11:44:21 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 4.644 min
Delta R.T.: 0.000 min
Response: 9035502
Conc: 21.05 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 5.571 min
Delta R.T.: 0.000 min
Response: 22638276
Conc: 22.19 ng/ml




#4 Heptachlor

R.T.: 6.164 min
Delta R.T.: -0.024 min
Response: 417539
Conc: 0.69 ng/ml



#4 Heptachlor


R.T.: 0.000 min
Exp R.T.: 7.195 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 6.164 min
Delta R.T.: 0.009 min
Response: 417539
Conc: 1.43 ng/ml

Instrument :
ECD_L
ClientSampleId :




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 6.799 min
Response: 0
Conc: N.D.




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 7.091 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 8.039 min
Delta R.T.: 0.005 min
Response: 4009163
Conc: 3.55 ng/ml



#10 gamma-Chlordane


R.T.: 0.000 min
Exp R.T.: 7.369 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#10 gamma-Chlordane

R.T.: 8.297 min
Delta R.T.: -0.006 min
Response: 2866371
Conc: 2.55 ng/ml




#13 Dieldrin

R.T.: 7.782 min
Delta R.T.: 0.002 min
Response: 59765
Conc: 0.11 ng/ml



#13 Dieldrin


R.T.: 0.000 min
Exp R.T.: 8.728 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 8.092 min
Delta R.T.: 0.020 min
Response: 536267
Conc: 1.07 ng/ml

Instrument :
ECD_L
ClientSampleId :




#14 Endrin

R.T.: 0.000 min
Exp R.T. : 8.966 min
Response: 0
Conc: N.D.




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 8.382 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 9.220 min
Delta R.T.: 0.019 min
Response: 646282
Conc: 0.71 ng/ml



#25 Chlordane-3


R.T.: 0.000 min
Exp R.T.: 7.371 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#25 Chlordane-3

R.T.: 8.297 min
Delta R.T.: -0.010 min
Response: 2866371
Conc: 18.51 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.423 min
Delta R.T.: -0.002 min
Response: 12693057
Conc: 22.57 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.493 min
Delta R.T.: 0.000 min
Response: 20669387
Conc: 23.17 ng/ml