

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072118\
 Data File : PL038406.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jul 2018 13:06
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 7/23/2018 1:06:39 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 01:37:02 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
 Quant Title : GC Extractables
 QLast Update : Sat Jul 14 00:12:12 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.327	3.928	349.9E6	140.2E6	46.995	53.408
28)	SA Decachlor...	7.996	8.934	342.2E6	121.7E6	45.108	49.788
Target Compounds							
2)	A alpha-BHC	3.756	4.307	555.8E6	209.4E6	47.038	54.099
3)	MA gamma-BHC...	4.032	4.587	501.7E6	187.1E6	47.858	51.807
4)	MA Heptachlor	4.315	5.097	486.9E6	180.6E6	45.773	56.502
5)	MB Aldrin	4.554	5.399	491.6E6	169.5E6	46.963	54.937
6)	B beta-BHC	4.285	4.749	224.6E6	84931367	47.574	55.990
7)	B delta-BHC	4.480	4.963	525.2E6	185.3E6	48.386	55.865
8)	B Heptachlo...	4.991	5.779	455.3E6	154.9E6	48.295	54.506
9)	A Endosulfan I	5.322	6.132	423.1E6	134.3E6	48.913	52.228
10)	B gamma-Chl...	5.214	6.011	472.7E6	152.7E6	47.825	52.666
11)	B alpha-Chl...	5.270	6.085	454.5E6	149.9E6	48.018	52.294
12)	B 4,4'-DDE	5.439	6.241	462.2E6	139.8E6	48.209	52.263
13)	MA Dieldrin	5.560	6.385	460.1E6	142.3E6	47.877m	52.266
14)	MA Endrin	5.815	6.600	395.0E6	114.6E6	45.467	47.981
15)	B Endosulfa...	6.090	6.803	393.0E6	121.9E6	49.561	50.778
16)	A 4,4'-DDD	5.951	6.723	381.5E6	111.0E6	52.006	51.876
17)	MA 4,4'-DDT	6.184	7.021	325.5E6	99113824	43.820	46.488
18)	B Endrin al...	6.258	6.925	312.1E6	104.1E6	49.017	53.345
19)	B Endosulfa...	6.469	7.148	370.4E6	115.1E6	49.032	52.800
20)	A Methoxychlor	6.733	7.477	174.3E6	55648813	45.296	48.872
21)	B Endrin ke...	6.951	7.612	418.2E6	123.7E6	53.302m	51.898

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072118\
Data File : PL038406.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jul 2018 13:06
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
7/23/2018 1:06:39 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 01:37:02 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
Quant Title : GC Extractables
QLast Update : Sat Jul 14 00:12:12 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

