

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072223\
 Data File : PL084108.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jul 2023 17:20
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 20 21:46:44 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071223.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 14 00:03:54 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.476	2.688	99797019	50914957	49.901	45.846
2)	SA Decachlor...	9.024	7.871	63771206	48311542	48.992	43.480
Target Compounds							
2)	A alpha-BHC	3.933	3.193	148.9E6	76574434	50.817	47.625
3)	MA gamma-BHC...	4.268	3.524	141.6E6	73827652	50.102	48.549
4)	MA Heptachlor	4.860	3.865	140.2E6	64923770	51.600	55.091
5)	MB Aldrin	5.202	4.145	140.3E6	69629023	48.982	46.178
6)	B beta-BHC	4.467	3.825	62216105	32079528	51.853	48.601
7)	B delta-BHC	4.713	4.054	139.8E6	73955820	52.445	48.892
8)	B Heptachlo...	5.633	4.652	126.8E6	61545588	47.507	47.926
9)	A Endosulfan I	6.018	5.023	111.0E6	57968897	46.857	47.263
10)	B gamma-Chl...	5.889	4.905	124.8E6	63555518	48.329	47.709
11)	B alpha-Chl...	5.969	4.969	120.7E6	63472378	47.493	47.117
12)	B 4,4'-DDE	6.148	5.168	100.3E6	59031635	49.866	48.058
13)	MA Dieldrin	6.295	5.290	112.1E6	63708374	46.892	48.918
14)	MA Endrin	6.524	5.565	87935919	55207253	46.174	50.223
15)	B Endosulfa...	6.745	5.862	85129001	55486872	45.419	48.277
16)	A 4,4'-DDD	6.667	5.726	75976635	51151289	51.737	53.319
17)	MA 4,4'-DDT	6.982	5.978	70933438	49485300	44.478	54.332
18)	B Endrin al...	6.876	6.044	66462280	44091971	45.364	49.049
19)	B Endosulfa...	7.112	6.269	80083266	53331128	46.326	48.526
20)	A Methoxychlor	7.464	6.562	35054648	27596372	44.814	57.583 #
21)	B Endrin ke...	7.595	6.774	82734009	65796582	47.210	50.923
22)	Mirex	8.071	6.955	60609189	48615521	42.356	47.631

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072223\
Data File : PL084108.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jul 2023 17:20
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 20 21:46:44 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071223.M
Quant Title : GC Extractables
QLast Update : Fri Jul 14 00:03:54 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

