

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072319\
 Data File : PL050718.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2019 15:30
 Operator : SM\AJ
 Sample : PB121646BSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB121646BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 01:09:19 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 19 01:19:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.318	3.930	190.2E6	53565104	18.933	18.774
28)	SA Decachlor...	7.979	8.961	205.2E6	56279229	20.270	19.901
Target Compounds							
2)	A alpha-BHC	3.745	4.318	899.7E6	216.3E6	53.688	52.413
3)	MA gamma-BHC...	4.020	4.601	848.6E6	205.9E6	55.171	53.149
4)	MA Heptachlor	4.305	5.102	888.8E6	216.3E6	54.171	52.166
5)	MB Aldrin	4.542	5.405	855.9E6	202.3E6	55.249	53.688
6)	B beta-BHC	4.275	4.778	359.8E6	94378354	54.883	54.590
7)	B delta-BHC	4.468	4.991	828.7E6	186.5E6	50.504	48.219
8)	B Heptachlor...	4.979	5.792	813.1E6	193.8E6	55.036	50.675
9)	A Endosulfan I	5.309	6.147	781.0E6	181.6E6	53.699	52.820
10)	B gamma-Chl...	5.201	6.026	849.7E6	196.6E6	53.835	53.199
11)	B alpha-Chl...	5.257	6.097	824.5E6	195.0E6	54.041	53.209
12)	B 4,4'-DDE	5.424	6.258	830.9E6	195.6E6	55.448	54.276
13)	MA Dieldrin	5.548	6.403	842.8E6	190.0E6	54.388	53.656
14)	MA Endrin	5.800	6.621	716.7E6	143.9E6	48.687	46.384
15)	B Endosulfa...	6.075	6.832	742.3E6	167.1E6	53.043	52.565
16)	A 4,4'-DDD	5.935	6.748	672.1E6	149.5E6	56.087	56.139
17)	MA 4,4'-DDT	6.168	7.045	628.4E6	134.8E6	50.736	49.921
18)	B Endrin al...	6.242	6.957	620.4E6	140.5E6	53.775	51.699
19)	B Endosulfa...	6.454	7.180	654.2E6	146.9E6	51.349	50.873
20)	A Methoxychlor	6.713	7.507	292.6E6	71246038	46.904	48.557
21)	B Endrin ke...	6.935	7.648	751.4E6	165.4E6	54.935	57.200
22)	Mirex	7.108	8.097	469.7E6	117.1E6	47.906	50.966

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072319\
Data File : PL050718.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2019 15:30
Operator : SM\AJ
Sample : PB121646BSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB121646BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 01:09:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
Quant Title : GC Extractables
QLast Update : Fri Jul 19 01:19:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

