

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072320\
 Data File : PL060121.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2020 09:54
 Operator : DD\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 01:55:46 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071020.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 10 18:52:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.465	4.012	108.3E6	148.1E6	56.003	55.883
28)	SA Decachlor...	8.223	9.099	80769018	100.5E6	49.097	48.074
Target Compounds							
2)	A alpha-BHC	3.903	4.402	163.4E6	218.8E6	56.832	56.686
3)	MA gamma-BHC...	4.186	4.687	140.8E6	204.3E6	57.007	56.328
4)	MA Heptachlor	4.489	5.198	126.5E6	188.7E6	56.409	56.093
5)	MB Aldrin	4.737	5.504	109.1E6	172.3E6	55.308	55.311
6)	B beta-BHC	4.437	4.857	55989326	85010369	54.824	56.230
7)	B delta-BHC	4.641	5.074	125.6E6	178.5E6	56.511	57.302
8)	B Heptachlo...	5.179	5.891	101.1E6	159.0E6	55.105	59.045
9)	A Endosulfan I	5.519	6.249	92178331	145.4E6	54.794	54.499
10)	B gamma-Chl...	5.407	6.126	99869654	158.1E6	55.050	55.211
11)	B alpha-Chl...	5.466	6.198	98619871	157.6E6	54.689	55.126
12)	B 4,4'-DDE	5.631	6.355	90832504	145.0E6	54.543	54.187
13)	MA Dieldrin	5.762	6.506	96098276	150.4E6	54.461	54.380
14)	MA Endrin	6.019	6.724	83474546	122.1E6	57.348	57.365
15)	B Endosulfa...	6.292	6.933	88248633	132.8E6	52.839	52.872
16)	A 4,4'-DDD	6.146	6.846	77949415	122.6E6	51.063	53.423
17)	MA 4,4'-DDT	6.387	7.146	74923934	112.9E6	57.626	57.323
18)	B Endrin al...	6.460	7.058	74349283	113.6E6	52.117	52.296
19)	B Endosulfa...	6.673	7.281	84212165	123.1E6	51.230	52.614
20)	A Methoxychlor	6.934	7.606	42684276	63206339	57.066	55.677
21)	B Endrin ke...	7.163	7.753	96331905	129.9E6	51.329	50.293
22)	Mirex	7.353	8.209	73315032	93309358	51.533	50.486

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072320\
Data File : PL060121.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2020 09:54
Operator : DD\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 01:55:46 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071020.M
Quant Title : GC Extractables
QLast Update : Fri Jul 10 18:52:06 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

