

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072420\
 Data File : PL060194.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jul 2020 18:13
 Operator : DD\AJ
 Sample : L3409-02
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 354

Manual Integrations
 APPROVED

Ankita
 7/27/2020 12:27:55 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 25 04:30:51 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071020.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 10 18:52:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.466	4.013	29143245	39965361	15.067	15.080
28)	SA Decachlor...	8.220	9.097	18923477	28150894	11.503m	13.460m
Target Compounds							
10)	B gamma-Chl...	5.406	6.126	694.3E6	1032.5E6	382.708m	360.562
11)	B alpha-Chl...	5.465	6.199	1350.7E6	2333.3E6	749.031	816.062
12)	B 4,4'-DDE	5.632	6.351	67450525	519.5E6	40.503m	194.088 #
17)	MA 4,4'-DDT	6.386	7.146	204.9E6	314.2E6	157.622	159.594

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072420\
Data File : PL060194.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2020 18:13
Operator : DD\AJ
Sample : L3409-02
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampled :
354


Manual Integrations
APPROVED

Ankita
7/27/2020 12:27:55 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 25 04:30:51 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071020.M
Quant Title : GC Extractables
QLast Update : Fri Jul 10 18:52:06 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.466 min
Delta R.T.: 0.000 min
Response: 29143245
Conc: 15.07 ng/ml

Instrument :
ECD_L
ClientSampleId :
354


Manual Integrations
APPROVED

Ankita
7/27/2020 12:27:55 PM




#1 Tetrachloro-m-xylene

R.T.: 4.013 min
Delta R.T.: 0.000 min
Response: 39965361
Conc: 15.08 ng/ml




#10 gamma-Chlordane

R.T.: 5.406 min
Delta R.T.: -0.004 min
Response: 694289073
Conc: 382.71 ng/ml m



#10 gamma-Chlordane

R.T.: 6.126 min
Delta R.T.: -0.002 min
Response: 1032500062
Conc: 360.56 ng/ml




#11 alpha-Chlordane

R.T.: 5.465 min
Delta R.T.: -0.003 min
Response: 1350725973
Conc: 749.03 ng/ml

Instrument :
ECD_L
ClientSampleId :
354


Manual Integrations
APPROVED

Ankita
7/27/2020 12:27:55 PM




#11 alpha-Chlordane

R.T.: 6.199 min
Delta R.T.: -0.001 min
Response: 2333329802
Conc: 816.06 ng/ml




#12 4,4'-DDE

R.T.: 5.632 min
Delta R.T.: -0.002 min
Response: 67450525
Conc: 40.50 ng/ml m



#12 4,4'-DDE

R.T.: 6.351 min
Delta R.T.: -0.007 min
Response: 519510900
Conc: 194.09 ng/ml




#17 4,4'-DDT

R.T.: 6.386 min
Delta R.T.: -0.006 min
Response: 204935946
Conc: 157.62 ng/ml

Instrument :
ECD_L
ClientSampleId :
354


Manual Integrations
APPROVED

Ankita
7/27/2020 12:27:55 PM




#17 4,4'-DDT

R.T.: 7.146 min
Delta R.T.: -0.003 min
Response: 314249061
Conc: 159.59 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.220 min
Delta R.T.: -0.011 min
Response: 18923477
Conc: 11.50 ng/ml m



#28 Decachlorobiphenyl

R.T.: 9.097 min
Delta R.T.: -0.006 min
Response: 28150894
Conc: 13.46 ng/ml m