

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072423\
 Data File : PL084152.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jul 2023 11:01
 Operator : AR\AJ
 Sample : 03670-01RE (Sig #1); 3670-01RE (Sig #2)
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 N48909RE

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 07/25/2023
 Supervised By : Ankita Jodhani 07/25/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 14:40:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071223.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 14 00:03:54 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.482	2.692	16222630	12470361	8.112m	11.229m#
28) SA Decachlor...	9.029	7.873	19671397	11387183	15.113	10.248 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072423\
Data File : PL084152.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2023 11:01
Operator : AR\AJ
Sample : 03670-01RE (Sig #1); 3670-01RE (Sig #2)
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
N48909RE


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 07/25/2023
Supervised By : Ankita Jodhani 07/25/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 14:40:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071223.M
Quant Title : GC Extractables
QLast Update : Fri Jul 14 00:03:54 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm





#1 Tetrachloro-m-xylene

R.T.: 3.482 min
Delta R.T.: 0.009 min
Response: 16222630
Conc: 8.11 ng/ml


Instrument :

ECD_L

ClientSampleId :
N48909RE


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/25/2023
Supervised By :Ankita Jodhani 07/25/2023




#1 Tetrachloro-m-xylene

R.T.: 2.692 min
Delta R.T.: 0.003 min
Response: 12470361
Conc: 11.23 ng/ml m



#28 Decachlorobiphenyl

R.T.: 9.029 min
Delta R.T.: 0.011 min
Response: 19671397
Conc: 15.11 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.873 min
Delta R.T.: 0.001 min
Response: 11387183
Conc: 10.25 ng/ml