

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072519\
 Data File : PL050779.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jul 2019 17:05
 Operator : SM\AJ
 Sample : K3990-02MS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-1-BMS

Manual Integrations
 APPROVED

Ankita
 7/26/2019 1:42:28 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 26 00:51:47 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 19 01:19:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.319	3.929	197.8E6	54845158	19.689	19.222
28)	SA Decachlor...	7.979	8.959	167.5E6	46136562	16.545	16.315
Target Compounds							
2)	A alpha-BHC	3.746	4.317	946.4E6	232.3E6	56.475	56.300
3)	MA gamma-BHC...	4.022	4.600	881.8E6	216.9E6	57.326	55.989
4)	MA Heptachlor	4.305	5.102	926.9E6	225.8E6	56.493	54.462
5)	MB Aldrin	4.543	5.404	870.0E6	203.0E6	56.157	53.863
6)	B beta-BHC	4.276	4.777	369.9E6	99712721	56.420	57.675
7)	B delta-BHC	4.469	4.990	842.3E6	178.4E6	51.333	46.129
8)	B Heptachlor...	4.979	5.791	810.2E6	191.5E6	54.842	50.070
9)	A Endosulfan I	5.310	6.146	768.2E6	181.5E6	52.818	52.788
10)	B gamma-Chl...	5.201	6.024	869.4E6	189.7E6	55.082	51.342
11)	B alpha-Chl...	5.257	6.096	821.7E6	193.3E6	53.858	52.731
12)	B 4,4'-DDE	5.425	6.257	781.7E6	190.7E6	52.159	52.911
13)	MA Dieldrin	5.548	6.402	808.6E6	189.6E6	52.176	53.546
14)	MA Endrin	5.801	6.620	773.4E6	159.3E6	52.541	51.353
15)	B Endosulfa...	6.076	6.831	699.0E6	165.9E6	49.949	52.172
16)	A 4,4'-DDD	5.935	6.747	611.0E6	141.5E6	50.982	53.150
17)	MA 4,4'-DDT	6.169	7.043	602.5E6	135.5E6	48.644	50.198
18)	B Endrin al...	6.243	6.953	536.0E6	129.7E6	46.462	47.739m
19)	B Endosulfa...	6.454	7.178	591.6E6	142.3E6	46.437	49.264
20)	A Methoxychlor	6.714	7.504	290.6E6	74440892	46.580	50.734
21)	B Endrin ke...	6.935	7.646	679.9E6	155.1E6	49.712	53.629
22)	Mirex	7.108	8.095	438.3E6	109.7E6	44.706	47.751

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072519\
 Data File : PL050779.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jul 2019 17:05
 Operator : SM\AJ
 Sample : K3990-02MS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampled :
 TP-1-BMS

Manual Integrations
 APPROVED

Ankita
 7/26/2019 1:42:28 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 26 00:51:47 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 19 01:19:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

