

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072619\
 Data File : PL050807.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jul 2019 12:56
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 26 22:47:07 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 19 01:19:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.319	3.928	519.2E6	138.4E6	51.674	48.491
28)	SA Decachlor...	7.977	8.957	548.5E6	140.8E6	54.173	49.796
Target Compounds							
2)	A alpha-BHC	3.746	4.317	866.4E6	204.8E6	51.703	49.634
3)	MA gamma-BHC...	4.021	4.600	805.7E6	192.9E6	52.381	49.807
4)	MA Heptachlor	4.304	5.101	833.3E6	200.0E6	50.784	48.222
5)	MB Aldrin	4.542	5.403	803.3E6	188.1E6	51.855	49.905
6)	B beta-BHC	4.275	4.776	338.0E6	89501676	51.558	51.769
7)	B delta-BHC	4.468	4.989	852.3E6	194.7E6	51.940	50.333
8)	B Heptachlo...	4.978	5.790	752.6E6	179.0E6	50.942	46.795
9)	A Endosulfan I	5.309	6.145	722.7E6	169.2E6	49.691	49.193
10)	B gamma-Chl...	5.201	6.024	793.8E6	183.9E6	50.297	49.761
11)	B alpha-Chl...	5.257	6.095	766.6E6	181.8E6	50.247	49.609
12)	B 4,4'-DDE	5.424	6.256	745.7E6	180.8E6	49.762	50.167
13)	MA Dieldrin	5.548	6.401	779.8E6	176.4E6	50.317	49.822
14)	MA Endrin	5.800	6.618	670.6E6	140.1E6	45.553	45.165
15)	B Endosulfa...	6.075	6.830	685.2E6	156.6E6	48.964	49.259
16)	A 4,4'-DDD	5.934	6.746	636.0E6	137.8E6	53.070	51.743
17)	MA 4,4'-DDT	6.168	7.043	567.7E6	125.7E6	45.834	46.567
18)	B Endrin al...	6.241	6.954	567.6E6	126.7E6	49.204	46.633
19)	B Endosulfa...	6.453	7.177	624.0E6	143.9E6	48.985	49.839
20)	A Methoxychlor	6.713	7.503	283.2E6	67813610	45.396	46.218
21)	B Endrin ke...	6.934	7.645	712.0E6	155.6E6	52.053	53.807
22)	Mirex	7.107	8.094	480.6E6	115.9E6	49.022	50.478
24)	Chlordane-2	0.000	5.403f	0	188.1E6	N.D.	1601.142 #
25)	Chlordane-3	5.201	6.024	793.8E6	183.9E6	567.116	382.303 #
26)	Chlordane-4	5.257	6.095	766.6E6	181.8E6	581.181	312.994 #
27)	Chlordane-5	6.075	0.000	685.2E6	0	1349.070	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072619\
Data File : PL050807.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jul 2019 12:56
Operator : SM\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 26 22:47:07 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
Quant Title : GC Extractables
QLast Update : Fri Jul 19 01:19:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.319 min
Delta R.T.: 0.000 min
Response: 519207684
Conc: 51.67 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 3.928 min
Delta R.T.: -0.002 min
Response: 138355811
Conc: 48.49 ng/ml




#2 alpha-BHC

R.T.: 3.746 min
Delta R.T.: 0.000 min
Response: 866431818
Conc: 51.70 ng/ml



#2 alpha-BHC


R.T.: 4.317 min
Delta R.T.: -0.002 min
Response: 204788192
Conc: 49.63 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.021 min
Delta R.T.: 0.000 min
Response: 805709678
Conc: 52.38 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 4.600 min
Delta R.T.: -0.002 min
Response: 192913054
Conc: 49.81 ng/ml




#4 Heptachlor

R.T.: 4.304 min
Delta R.T.: 0.000 min
Response: 833262945
Conc: 50.78 ng/ml



#4 Heptachlor


R.T.: 5.101 min
Delta R.T.: -0.003 min
Response: 199963628
Conc: 48.22 ng/ml



#5 Aldrin


R.T.: 4.542 min
Delta R.T.: -0.001 min
Response: 803314811
Conc: 51.86 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 5.403 min
Delta R.T.: -0.002 min
Response: 188058102
Conc: 49.90 ng/ml




#6 beta-BHC

R.T.: 4.275 min
Delta R.T.: 0.000 min
Response: 338004266
Conc: 51.56 ng/ml



#6 beta-BHC


R.T.: 4.776 min
Delta R.T.: -0.002 min
Response: 89501676
Conc: 51.77 ng/ml



#7 delta-BHC


R.T.: 4.468 min
Delta R.T.: 0.000 min
Response: 852269313
Conc: 51.94 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 4.989 min
Delta R.T.: -0.002 min
Response: 194667868
Conc: 50.33 ng/ml




#8 Heptachlor epoxide

R.T.: 4.978 min
Delta R.T.: -0.001 min
Response: 752598645
Conc: 50.94 ng/ml



#8 Heptachlor epoxide


R.T.: 5.790 min
Delta R.T.: -0.003 min
Response: 178984176
Conc: 46.80 ng/ml



#9 Endosulfan I


R.T.: 5.309 min
Delta R.T.: -0.001 min
Response: 722695660
Conc: 49.69 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 6.145 min
Delta R.T.: -0.003 min
Response: 169168967
Conc: 49.19 ng/ml




#10 gamma-Chlordane

R.T.: 5.201 min
Delta R.T.: 0.000 min
Response: 793841885
Conc: 50.30 ng/ml



#10 gamma-Chlordane


R.T.: 6.024 min
Delta R.T.: -0.002 min
Response: 183880125
Conc: 49.76 ng/ml



#11 alpha-Chlordane


R.T.: 5.257 min
Delta R.T.: 0.000 min
Response: 766566293
Conc: 50.25 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 6.095 min
Delta R.T.: -0.003 min
Response: 181823949
Conc: 49.61 ng/ml




#12 4,4'-DDE

R.T.: 5.424 min
Delta R.T.: -0.001 min
Response: 745723068
Conc: 49.76 ng/ml



#12 4,4'-DDE


R.T.: 6.256 min
Delta R.T.: -0.002 min
Response: 180772541
Conc: 50.17 ng/ml



#13 Dieldrin


R.T.: 5.548 min
Delta R.T.: 0.000 min
Response: 779750150
Conc: 50.32 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 6.401 min
Delta R.T.: -0.003 min
Response: 176427189
Conc: 49.82 ng/ml




#14 Endrin

R.T.: 5.800 min
Delta R.T.: -0.002 min
Response: 670557298
Conc: 45.55 ng/ml



#14 Endrin


R.T.: 6.618 min
Delta R.T.: -0.003 min
Response: 140076698
Conc: 45.17 ng/ml



#15 Endosulfan II


R.T.: 6.075 min
Delta R.T.: 0.000 min
Response: 685242737
Conc: 48.96 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 6.830 min
Delta R.T.: -0.003 min
Response: 156602995
Conc: 49.26 ng/ml




#16 4,4'-DDD

R.T.: 5.934 min
Delta R.T.: -0.001 min
Response: 635987244
Conc: 53.07 ng/ml



#16 4,4'-DDD


R.T.: 6.746 min
Delta R.T.: -0.003 min
Response: 137785183
Conc: 51.74 ng/ml



#17 4,4'-DDT


R.T.: 6.168 min
Delta R.T.: -0.001 min
Response: 567690295
Conc: 45.83 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 7.043 min
Delta R.T.: -0.003 min
Response: 125711098
Conc: 46.57 ng/ml




#18 Endrin aldehyde

R.T.: 6.241 min
Delta R.T.: -0.002 min
Response: 567641091
Conc: 49.20 ng/ml



#18 Endrin aldehyde


R.T.: 6.954 min
Delta R.T.: -0.003 min
Response: 126730722
Conc: 46.63 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.453 min
Delta R.T.: -0.002 min
Response: 624041265
Conc: 48.98 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 7.177 min
Delta R.T.: -0.003 min
Response: 143947242
Conc: 49.84 ng/ml




#20 Methoxychlor

R.T.: 6.713 min
Delta R.T.: -0.001 min
Response: 283225450
Conc: 45.40 ng/ml



#20 Methoxychlor


R.T.: 7.503 min
Delta R.T.: -0.004 min
Response: 67813610
Conc: 46.22 ng/ml



#21 Endrin ketone


R.T.: 6.934 min
Delta R.T.: -0.002 min
Response: 711952946
Conc: 52.05 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 7.645 min
Delta R.T.: -0.003 min
Response: 155570986
Conc: 53.81 ng/ml




#22 Mirex

R.T.: 7.107 min
Delta R.T.: -0.002 min
Response: 480621654
Conc: 49.02 ng/ml



#22 Mirex


R.T.: 8.094 min
Delta R.T.: -0.004 min
Response: 115948375
Conc: 50.48 ng/ml



#24 Chlordane-2


R.T.: 0.000 min
Exp R.T.: 4.647 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#24 Chlordane-2

R.T.: 5.403 min
Delta R.T.: 0.020 min
Response: 188058102
Conc: 1601.14 ng/ml




#25 Chlordane-3

R.T.: 5.201 min
Delta R.T.: -0.002 min
Response: 793841885
Conc: 567.12 ng/ml



#25 Chlordane-3


R.T.: 6.024 min
Delta R.T.: -0.003 min
Response: 183880125
Conc: 382.30 ng/ml



#26 Chlordane-4


R.T.: 5.257 min
Delta R.T.: 0.000 min
Response: 766566293
Conc: 581.18 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#26 Chlordane-4

R.T.: 6.095 min
Delta R.T.: -0.005 min
Response: 181823949
Conc: 312.99 ng/ml




#27 Chlordane-5

R.T.: 6.075 min
Delta R.T.: -0.004 min
Response: 685242737
Conc: 1349.07 ng/ml



#27 Chlordane-5


R.T.: 0.000 min
Exp R.T.: 6.902 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 7.977 min
Delta R.T.: -0.004 min
Response: 548519876
Conc: 54.17 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 8.957 min
Delta R.T.: -0.005 min
Response: 140819612
Conc: 49.80 ng/ml