

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072720\
 Data File : PL060264.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jul 2020 21:47
 Operator : DD\AJ
 Sample : L3436-10 10X
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 10

Manual Integrations
 APPROVED

Ankita
 7/29/2020 4:34:27 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 28 01:13:33 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071020.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 10 18:52:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.465	4.013	2279885	2470575	1.179	0.932
28)	SA Decachlor...	8.225	9.095	2422061	2714458	1.472	1.298m
Target Compounds							
10)	B gamma-Chl...	5.408	6.124	2684058	2980531	1.480m	1.041m#
11)	B alpha-Chl...	5.466	6.200	2607910	3244929	1.446m	1.135m
16)	A 4,4'-DDD	6.149	6.847	827048	2569946	0.542m	1.120 #
17)	MA 4,4'-DDT	6.388	7.147	1289376	2367273	0.992	1.202m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072720\
Data File : PL060264.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jul 2020 21:47
Operator : DD\AJ
Sample : L3436-10 10X
Misc :
ALS Vial : 37 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampled :
10


Manual Integrations
APPROVED

Ankita
7/29/2020 4:34:27 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 28 01:13:33 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071020.M
Quant Title : GC Extractables
QLast Update : Fri Jul 10 18:52:06 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.465 min
Delta R.T.: 0.000 min
Response: 2279885
Conc: 1.18 ng/ml

Instrument :
ECD_L
ClientSampleId :
10


Manual Integrations
APPROVED

Ankita
7/29/2020 4:34:27 PM




#1 Tetrachloro-m-xylene

R.T.: 4.013 min
Delta R.T.: 0.000 min
Response: 2470575
Conc: 0.93 ng/ml




#10 gamma-Chlordane

R.T.: 5.408 min
Delta R.T.: -0.002 min
Response: 2684058
Conc: 1.48 ng/ml m



#10 gamma-Chlordane

R.T.: 6.124 min
Delta R.T.: -0.003 min
Response: 2980531
Conc: 1.04 ng/ml m




#11 alpha-Chlordane

R.T.: 5.466 min
Delta R.T.: -0.002 min
Response: 2607910
Conc: 1.45 ng/ml m

Instrument :
ECD_L
ClientSampleId :
10


Manual Integrations
APPROVED

Ankita
7/29/2020 4:34:27 PM




#11 alpha-Chlordane

R.T.: 6.200 min
Delta R.T.: 0.000 min
Response: 3244929
Conc: 1.13 ng/ml m




#16 4,4'-DDD

R.T.: 6.149 min
Delta R.T.: -0.001 min
Response: 827048
Conc: 0.54 ng/ml m



#16 4,4'-DDD

R.T.: 6.847 min
Delta R.T.: -0.002 min
Response: 2569946
Conc: 1.12 ng/ml




#17 4,4'-DDT

R.T.: 6.388 min
Delta R.T.: -0.004 min
Response: 1289376
Conc: 0.99 ng/ml

Instrument :
ECD_L
ClientSampleId :
10


Manual Integrations
APPROVED

Ankita
7/29/2020 4:34:27 PM




#17 4,4'-DDT

R.T.: 7.147 min
Delta R.T.: -0.002 min
Response: 2367273
Conc: 1.20 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.225 min
Delta R.T.: -0.006 min
Response: 2422061
Conc: 1.47 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.095 min
Delta R.T.: -0.007 min
Response: 2714458
Conc: 1.30 ng/ml m