

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072823\
 Data File : PL084336.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2023 13:32
 Operator : AR\AJ
 Sample : 03795-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TOPSOIL-OU3-DENALI-TEMP

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 07/31/2023
 Supervised By : Ankita Jodhani 07/31/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 28 22:31:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071223.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 14 00:03:54 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.477	2.689	19021602	9007451	9.511	8.111
28) SA Decachlor...	9.027	7.870	16875227	13090751	12.964	11.781
Target Compounds						
10) B gamma-Chl...	5.890	4.906	4234069	849567	1.640m	0.638 #
11) B alpha-Chl...	5.975	4.967	11425944	2907118	4.495	2.158 #
12) B 4,4'-DDE	6.149	5.167	19652570	11412812	9.768	9.291
16) A 4,4'-DDD	6.671	5.724	3133772	1066101	2.134m	1.111 #
17) MA 4,4'-DDT	6.982	5.976	7754607	5382673	4.862m	5.910

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL072823\
Data File : PL084336.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2023 13:32
Operator : AR\AJ
Sample : 03795-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TOPSOIL-OU3-DENALI-TEMP

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 07/31/2023
Supervised By : Ankita Jodhani 07/31/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 28 22:31:34 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071223.M
Quant Title : GC Extractables
QLast Update : Fri Jul 14 00:03:54 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

