

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL073019\
 Data File : PL050922.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jul 2019 04:42
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 7/31/2019 3:30:41 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 31 04:50:46 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 19 01:19:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.318	3.929	537.3E6	142.1E6	53.477	49.817
28)	SA Decachlor...	7.978	8.959	533.5E6	139.3E6	52.692	49.242
Target Compounds							
2)	A alpha-BHC	3.746	4.317	903.4E6	211.2E6	53.911	51.195
3)	MA gamma-BHC...	4.021	4.601	839.1E6	199.1E6	54.553	51.409
4)	MA Heptachlor	4.305	5.102	888.6E6	210.5E6	54.159	50.764
5)	MB Aldrin	4.542	5.404	835.9E6	192.5E6	53.958	51.096
6)	B beta-BHC	4.276	4.777	351.6E6	92060780	53.632	53.249
7)	B delta-BHC	4.468	4.990	885.4E6	197.9E6	53.958	51.163
8)	B Heptachlo...	4.979	5.791	782.6E6	181.7E6	52.974	47.501
9)	A Endosulfan I	5.309	6.145	753.0E6	171.0E6	51.771	49.714
10)	B gamma-Chl...	5.201	6.025	824.5E6	185.1E6	52.236	50.089
11)	B alpha-Chl...	5.257	6.096	797.3E6	183.7E6	52.262	50.112
12)	B 4,4'-DDE	5.424	6.257	771.0E6	181.9E6	51.451	50.484
13)	MA Dieldrin	5.548	6.402	806.5E6	178.3E6	52.046	50.357
14)	MA Endrin	5.800	6.619	740.4E6	156.0E6	50.296	50.305
15)	B Endosulfa...	6.075	6.831	686.3E6	157.2E6	49.037	49.462
16)	A 4,4'-DDD	5.935	6.747	610.0E6	135.3E6	50.898	50.803
17)	MA 4,4'-DDT	6.168	7.043	629.5E6	140.0E6	50.826	51.845
18)	B Endrin al...	6.241	6.954	558.0E6	132.8E6	48.367	48.869m
19)	B Endosulfa...	6.453	7.178	615.2E6	144.5E6	48.294	50.040
20)	A Methoxychlor	6.714	7.504	309.4E6	74375502	49.598	50.690
21)	B Endrin ke...	6.934	7.646	663.3E6	147.7E6	48.494	51.083
22)	Mirex	7.107	8.095	456.3E6	114.7E6	46.537	49.928

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL073019\
Data File : PL050922.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jul 2019 04:42
Operator : SM\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
7/31/2019 3:30:41 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 31 04:50:46 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
Quant Title : GC Extractables
QLast Update : Fri Jul 19 01:19:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

