

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080118\
 Data File : PL038747.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2018 06:08
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 8/2/2018 5:38:44 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 07:51:29 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
 Quant Title : GC Extractables
 QLast Update : Sat Jul 14 00:12:12 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.325	3.928	320.1E6	127.3E6	42.984	48.500
28)	SA Decachlor...	7.992	8.934	304.7E6	119.8E6	40.165	49.027
Target Compounds							
2)	A alpha-BHC	3.754	4.307	532.8E6	203.3E6	45.094	52.540
3)	MA gamma-BHC...	4.029	4.587	484.7E6	181.6E6	46.237m	50.286
4)	MA Heptachlor	4.313	5.097	478.8E6	180.8E6	45.011	56.581 #
5)	MB Aldrin	4.551	5.399	483.5E6	170.0E6	46.189	55.079
6)	B beta-BHC	4.284	4.749	215.8E6	82036719	45.715	54.082
7)	B delta-BHC	4.478	4.963	510.9E6	181.4E6	47.073	54.683
8)	B Heptachlo...	4.989	5.779	390.6E6	152.6E6	41.427	53.690 #
9)	A Endosulfan I	5.319	6.131	424.4E6	132.7E6	49.066	51.614
10)	B gamma-Chl...	5.210	6.011	462.5E6	148.7E6	46.789	51.299
11)	B alpha-Chl...	5.267	6.084	445.7E6	148.0E6	47.089	51.641
12)	B 4,4'-DDE	5.434	6.241	429.6E6	141.3E6	44.809m	52.839
13)	MA Dieldrin	5.558	6.385	447.7E6	144.1E6	46.583	52.956
14)	MA Endrin	5.811	6.600	390.4E6	118.5E6	44.931	49.621
15)	B Endosulfa...	6.085	6.803	394.0E6	123.3E6	49.680m	51.330
16)	A 4,4'-DDD	5.947	6.722	355.0E6	110.1E6	48.399	51.440
17)	MA 4,4'-DDT	6.180	7.021	303.4E6	102.7E6	40.843m	48.180
18)	B Endrin al...	6.253	6.925	286.7E6	104.6E6	45.022m	53.609
19)	B Endosulfa...	6.464	7.148	354.0E6	117.0E6	46.862m	53.663
20)	A Methoxychlor	6.729	7.476	164.3E6	50192218	42.697	44.080
21)	B Endrin ke...	6.948	7.611	353.2E6	125.3E6	45.014	52.567m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080118\
Data File : PL038747.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2018 06:08
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
8/2/2018 5:38:44 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 07:51:29 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
Quant Title : GC Extractables
QLast Update : Sat Jul 14 00:12:12 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

