

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080118\
 Data File : PL038805.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Aug 2018 10:18
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 8/2/2018 5:41:16 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 10:29:57 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
 Quant Title : GC Extractables
 QLast Update : Sat Jul 14 00:12:12 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.325	3.928	339.7E6	133.8E6	45.614	50.969
2)	SA Decachlor...	7.990	8.935	304.4E6	115.0E6	40.122m	47.036
Target Compounds							
2)	A alpha-BHC	3.754	4.307	564.2E6	211.7E6	47.749	54.706
3)	MA gamma-BHC...	4.028	4.587	511.5E6	190.1E6	48.786m	52.636
4)	MA Heptachlor	4.312	5.096	500.0E6	187.3E6	47.003	58.596
5)	MB Aldrin	4.551	5.399	495.7E6	175.7E6	47.355	56.929
6)	B beta-BHC	4.284	4.749	226.6E6	86179198	48.005	56.813
7)	B delta-BHC	4.478	4.963	529.6E6	191.4E6	48.796	57.718
8)	B Heptachlo...	4.987	5.778	451.7E6	162.5E6	47.909m	57.191
9)	A Endosulfan I	5.319	6.131	422.1E6	137.6E6	48.800	53.507
10)	B gamma-Chl...	5.210	6.011	475.2E6	156.6E6	48.076	54.024
11)	B alpha-Chl...	5.266	6.085	454.4E6	154.2E6	48.003	53.795
12)	B 4,4'-DDE	5.435	6.240	477.3E6	145.2E6	49.778	54.290
13)	MA Dieldrin	5.558	6.385	461.3E6	145.0E6	48.002	53.286
14)	MA Endrin	5.811	6.600	405.4E6	125.3E6	46.658	52.458
15)	B Endosulfa...	6.085	6.803	401.3E6	123.1E6	50.610m	51.245
16)	A 4,4'-DDD	5.947	6.723	373.8E6	113.4E6	50.964	53.000
17)	MA 4,4'-DDT	6.181	7.021	333.5E6	103.5E6	44.895	48.561
18)	B Endrin al...	6.254	6.925	305.7E6	104.2E6	48.016	53.400
19)	B Endosulfa...	6.465	7.148	365.6E6	115.6E6	48.399	53.043
20)	A Methoxychlor	6.729	7.476	173.0E6	58268819	44.974	51.173
21)	B Endrin ke...	6.948	7.611	378.4E6	116.6E6	48.224	48.911

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080118\
Data File : PL038805.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Aug 2018 10:18
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
Client Sampled :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
8/2/2018 5:41:16 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 10:29:57 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
Quant Title : GC Extractables
QLast Update : Sat Jul 14 00:12:12 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

