

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080119\
 Data File : PL050962.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2019 17:41
 Operator : SM\AJ
 Sample : K3618-10MSD
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 HD-01-073019MSD

Manual Integrations
 APPROVED

Ankita
 8/2/2019 2:06:14 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 01:02:00 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 19 01:19:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.319	3.929	212.6E6	63042138	21.156	22.095
28)	SA Decachlor...	7.978	8.957	247.2E6	65632081	24.419	23.209
Target Compounds							
2)	A alpha-BHC	3.746	4.317	1101.2E6	283.6E6	65.712	68.745
3)	MA gamma-BHC...	4.021	4.600	1033.1E6	268.0E6	67.164	69.189
4)	MA Heptachlor	4.305	5.101	1125.1E6	290.7E6	68.570	70.097
5)	MB Aldrin	4.543	5.403	1066.3E6	269.0E6	68.833	71.383
6)	B beta-BHC	4.276	4.777	424.8E6	123.9E6	64.796	71.649
7)	B delta-BHC	4.469	4.990	1015.2E6	241.0E6	61.870	62.310
8)	B Heptachlo...	4.979	5.790	984.6E6	253.1E6	66.647	66.173
9)	A Endosulfan I	5.309	6.145	987.6E6	241.0E6	67.907	70.072
10)	B gamma-Chl...	5.201	6.024	1080.5E6	259.5E6	68.461	70.223
11)	B alpha-Chl...	5.257	6.095	1040.5E6	256.8E6	68.203	70.054
12)	B 4,4'-DDE	5.423	6.256	1008.0E6	257.7E6	67.261	71.527
13)	MA Dieldrin	5.548	6.401	1055.9E6	256.6E6	68.134	72.465
14)	MA Endrin	5.800	6.618	1012.1E6	225.6E6	68.755	72.729
15)	B Endosulfa...	6.075	6.830	965.9E6	227.1E6	69.020	71.444
16)	A 4,4'-DDD	5.935	6.746	813.6E6	194.2E6	67.890	72.913
17)	MA 4,4'-DDT	6.168	7.042	871.6E6	204.2E6	70.370	75.638
18)	B Endrin al...	6.242	6.953	754.9E6	186.9E6	65.438	68.775m
19)	B Endosulfa...	6.453	7.177	852.8E6	202.3E6	66.939	70.048
20)	A Methoxychlor	6.713	7.503	419.0E6	108.3E6	67.166	73.812
21)	B Endrin ke...	6.934	7.645	948.7E6	217.6E6	69.364	75.251
22)	Mirex	7.107	8.094	612.3E6	155.3E6	62.452	67.629

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080119\
Data File : PL050962.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2019 17:41
Operator : SM\AJ
Sample : K3618-10MSD
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
HD-01-073019MSD

Manual Integrations
APPROVED

Ankita
8/2/2019 2:06:14 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 01:02:00 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
Quant Title : GC Extractables
QLast Update : Fri Jul 19 01:19:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

