

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080120\
 Data File : PL060446.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jul 2020 20:36
 Operator : DD\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 00:05:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL073020.M
 Quant Title : GC Extractables
 QLast Update : Thu Jul 30 04:13:43 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.464	4.012	88260989	121.5E6	44.284	46.441
28)	SA Decachlor...	8.223	9.102	72234481	82764394	47.333	46.015
Target Compounds							
2)	A alpha-BHC	3.901	4.402	129.8E6	178.6E6	44.511	46.649
3)	MA gamma-BHC...	4.185	4.687	113.7E6	166.0E6	44.821	46.108
4)	MA Heptachlor	4.488	5.198	102.3E6	154.7E6	43.897	45.464
5)	MB Aldrin	4.736	5.504	90250928	141.7E6	45.531	46.184
6)	B beta-BHC	4.436	4.857	45251313	69379814	44.618	46.139
7)	B delta-BHC	4.640	5.075	102.6E6	137.9E6	45.239	54.474
8)	B Heptachlo...	5.179	5.891	83109018	125.6E6	43.840	51.823
9)	A Endosulfan I	5.517	6.250	74637971	120.8E6	44.637	44.679
10)	B gamma-Chl...	5.406	6.127	81080917	130.0E6	44.751	45.992
11)	B alpha-Chl...	5.465	6.200	80638177	130.8E6	44.748	44.817
12)	B 4,4'-DDE	5.630	6.357	69557268	118.9E6	41.260	44.961
13)	MA Dieldrin	5.761	6.507	78929558	123.1E6	44.595	44.892
14)	MA Endrin	6.018	6.726	71605676	103.3E6	42.957	44.095
15)	B Endosulfa...	6.291	6.935	71898514	106.4E6	42.351	44.172
16)	A 4,4'-DDD	6.145	6.849	66188840	100.2E6	45.809	48.123
17)	MA 4,4'-DDT	6.386	7.149	55878123	85499647	40.906	42.276
18)	B Endrin al...	6.459	7.060	61310541	92838390	42.678	46.334
19)	B Endosulfa...	6.671	7.283	69769828	101.6E6	43.785	46.202
20)	A Methoxychlor	6.933	7.609	34588555	49381394	42.280	42.881
21)	B Endrin ke...	7.161	7.756	78757242	104.1E6	46.713	47.243
22)	Mirex	7.351	8.213	61232392	74223451	44.852	47.398

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080120\
Data File : PL060446.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jul 2020 20:36
Operator : DD\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 00:05:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL073020.M
Quant Title : GC Extractables
QLast Update : Thu Jul 30 04:13:43 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

