

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080123\
 Data File : PL084413.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2023 00:07
 Operator : AR\AJ
 Sample : 03834-01
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 B-P2A

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 05:37:17 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL080123.M
 Quant Title : GC Extractables
 QLast Update : Mon Jul 31 17:26:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.476	2.686	21607135	10140323	9.352	9.275
28)	SA Decachlor...	9.024	7.868	14326770	11051742	9.329	9.038
Target Compounds							
11)	B alpha-Chl...	5.975	4.964	2899821	580068	1.019	0.412 #
12)	B 4,4'-DDE	6.147	5.164	1234581	585859	0.540	0.465



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080123\
Data File : PL084413.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2023 00:07
Operator : AR\AJ
Sample : 03834-01
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
B-P2A

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 05:37:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL080123.M
Quant Title : GC Extractables
QLast Update : Mon Jul 31 17:26:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.476 min
Delta R.T.: 0.000 min
Response: 21607135
Conc: 9.35 ng/ml

Instrument :
ECD_L
ClientSampleId :
B-P2A




#1 Tetrachloro-m-xylene

R.T.: 2.686 min
Delta R.T.: 0.000 min
Response: 10140323
Conc: 9.27 ng/ml




#11 alpha-Chlordane

R.T.: 5.975 min
Delta R.T.: 0.005 min
Response: 2899821
Conc: 1.02 ng/ml



#11 alpha-Chlordane


R.T.: 4.964 min
Delta R.T.: -0.004 min
Response: 580068
Conc: 0.41 ng/ml



#12 4,4'-DDE


R.T.: 6.147 min
Delta R.T.: -0.002 min
Response: 1234581
Conc: 0.54 ng/ml

Instrument :
ECD_L
ClientSampleId :
B-P2A




#12 4,4'-DDE

R.T.: 5.164 min
Delta R.T.: -0.002 min
Response: 585859
Conc: 0.47 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.024 min
Delta R.T.: -0.002 min
Response: 14326770
Conc: 9.33 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.868 min
Delta R.T.: 0.000 min
Response: 11051742
Conc: 9.04 ng/ml