

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080218\
 Data File : PL038856.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Aug 2018 00:51
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 03 03:58:19 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
 Quant Title : GC Extractables
 QLast Update : Sat Jul 14 00:12:12 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.325	3.928	312.2E6	129.2E6	41.925	49.235
28)	SA Decachlor...	7.991	8.935	286.0E6	109.9E6	37.701	44.967
Target Compounds							
2)	A alpha-BHC	3.754	4.307	527.1E6	204.3E6	44.612	52.781
3)	MA gamma-BHC...	4.030	4.587	464.0E6	182.6E6	44.256	50.566
4)	MA Heptachlor	4.313	5.096	462.1E6	178.2E6	43.442	55.752 #
5)	MB Aldrin	4.551	5.399	460.4E6	168.4E6	43.986	54.559
6)	B beta-BHC	4.285	4.749	212.2E6	83411595	44.952	54.989
7)	B delta-BHC	4.478	4.963	495.4E6	184.0E6	45.645	55.485
8)	B Heptachlo...	4.989	5.779	418.8E6	153.8E6	44.415	54.107
9)	A Endosulfan I	5.319	6.131	396.2E6	132.6E6	45.802	51.548
10)	B gamma-Chl...	5.211	6.011	443.5E6	150.7E6	44.871	51.969
11)	B alpha-Chl...	5.267	6.085	425.6E6	148.0E6	44.959	51.624
12)	B 4,4'-DDE	5.435	6.241	429.2E6	139.9E6	44.763	52.301
13)	MA Dieldrin	5.559	6.385	434.5E6	137.8E6	45.217	50.620
14)	MA Endrin	5.812	6.600	372.8E6	111.1E6	42.903	46.506
15)	B Endosulfa...	6.087	6.803	357.2E6	116.5E6	45.047	48.513
16)	A 4,4'-DDD	5.947	6.723	356.1E6	104.5E6	48.540	48.841
17)	MA 4,4'-DDT	6.181	7.021	306.6E6	94416320	41.271	44.285
18)	B Endrin al...	6.255	6.926	289.0E6	98419501	45.393	50.436
19)	B Endosulfa...	6.466	7.148	338.1E6	108.8E6	44.765	49.915
20)	A Methoxychlor	6.729	7.477	161.8E6	54420567	42.048	47.793
21)	B Endrin ke...	6.949	7.612	362.2E6	114.0E6	46.158	47.844
23)	Chlordane-1	0.000	4.963	0	184.0E6	N.D.	1587.966 #
24)	Chlordane-2	0.000	5.399	0	168.4E6	N.D.	1406.111 #
25)	Chlordane-3	5.211	6.011	443.5E6	150.7E6	381.234	369.378
26)	Chlordane-4	5.267	6.085	425.6E6	148.0E6	404.773	301.163 #
27)	Chlordane-5	5.947	6.926	356.1E6	98419501	2223.603	1098.683 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080218\
Data File : PL038856.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Aug 2018 00:51
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 03 03:58:19 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
Quant Title : GC Extractables
QLast Update : Sat Jul 14 00:12:12 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.325 min
Delta R.T.: -0.004 min
Response: 312194031
Conc: 41.93 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.928 min
Delta R.T.: -0.002 min
Response: 129239647
Conc: 49.24 ng/ml




#2 alpha-BHC

R.T.: 3.754 min
Delta R.T.: -0.004 min
Response: 527108347
Conc: 44.61 ng/ml



#2 alpha-BHC


R.T.: 4.307 min
Delta R.T.: -0.001 min
Response: 204271375
Conc: 52.78 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.030 min
Delta R.T.: -0.004 min
Response: 463969754
Conc: 44.26 ng/ml

Instrument :
ECD_L
ClientSampleId :




#3 gamma-BHC (Lindane)

R.T.: 4.587 min
Delta R.T.: -0.002 min
Response: 182635211
Conc: 50.57 ng/ml




#4 Heptachlor

R.T.: 4.313 min
Delta R.T.: -0.006 min
Response: 462106647
Conc: 43.44 ng/ml



#4 Heptachlor


R.T.: 5.096 min
Delta R.T.: -0.003 min
Response: 178197952
Conc: 55.75 ng/ml



#5 Aldrin


R.T.: 4.551 min
Delta R.T.: -0.007 min
Response: 460411485
Conc: 43.99 ng/ml

Instrument :
ECD_L
ClientSampleId :




#5 Aldrin

R.T.: 5.399 min
Delta R.T.: -0.003 min
Response: 168381953
Conc: 54.56 ng/ml




#6 beta-BHC

R.T.: 4.285 min
Delta R.T.: -0.005 min
Response: 212186713
Conc: 44.95 ng/ml



#6 beta-BHC


R.T.: 4.749 min
Delta R.T.: 0.000 min
Response: 83411595
Conc: 54.99 ng/ml



#7 delta-BHC


R.T.: 4.478 min
Delta R.T.: -0.006 min
Response: 495433726
Conc: 45.64 ng/ml

Instrument :
ECD_L
ClientSampleId :




#7 delta-BHC

R.T.: 4.963 min
Delta R.T.: 0.000 min
Response: 184021365
Conc: 55.48 ng/ml




#8 Heptachlor epoxide

R.T.: 4.989 min
Delta R.T.: -0.008 min
Response: 418765886
Conc: 44.42 ng/ml



#8 Heptachlor epoxide


R.T.: 5.779 min
Delta R.T.: -0.002 min
Response: 153766941
Conc: 54.11 ng/ml



#9 Endosulfan I


R.T.: 5.319 min
Delta R.T.: -0.008 min
Response: 396172548
Conc: 45.80 ng/ml

Instrument :
ECD_L
ClientSampleId :




#9 Endosulfan I

R.T.: 6.131 min
Delta R.T.: -0.002 min
Response: 132568974
Conc: 51.55 ng/ml




#10 gamma-Chlordane

R.T.: 5.211 min
Delta R.T.: -0.008 min
Response: 443501563
Conc: 44.87 ng/ml



#10 gamma-Chlordane


R.T.: 6.011 min
Delta R.T.: -0.002 min
Response: 150669315
Conc: 51.97 ng/ml



#11 alpha-Chlordane


R.T.: 5.267 min
Delta R.T.: -0.008 min
Response: 425570510
Conc: 44.96 ng/ml

Instrument :
ECD_L
ClientSampleId :




#11 alpha-Chlordane

R.T.: 6.085 min
Delta R.T.: -0.002 min
Response: 147965496
Conc: 51.62 ng/ml




#12 4,4'-DDE

R.T.: 5.435 min
Delta R.T.: -0.009 min
Response: 429182965
Conc: 44.76 ng/ml



#12 4,4'-DDE


R.T.: 6.241 min
Delta R.T.: -0.002 min
Response: 139857899
Conc: 52.30 ng/ml



#13 Dieldrin


R.T.: 5.559 min
Delta R.T.: -0.008 min
Response: 434543990
Conc: 45.22 ng/ml

Instrument :
ECD_L
ClientSampleId :




#13 Dieldrin

R.T.: 6.385 min
Delta R.T.: -0.002 min
Response: 137776485
Conc: 50.62 ng/ml




#14 Endrin

R.T.: 5.812 min
Delta R.T.: -0.009 min
Response: 372756984
Conc: 42.90 ng/ml



#14 Endrin


R.T.: 6.600 min
Delta R.T.: -0.002 min
Response: 111062864
Conc: 46.51 ng/ml



#15 Endosulfan II


R.T.: 6.087 min
Delta R.T.: -0.009 min
Response: 357231032
Conc: 45.05 ng/ml

Instrument :
ECD_L
ClientSampleId :




#15 Endosulfan II

R.T.: 6.803 min
Delta R.T.: -0.002 min
Response: 116500099
Conc: 48.51 ng/ml




#16 4,4'-DDD

R.T.: 5.947 min
Delta R.T.: -0.009 min
Response: 356072588
Conc: 48.54 ng/ml



#16 4,4'-DDD


R.T.: 6.723 min
Delta R.T.: -0.002 min
Response: 104540473
Conc: 48.84 ng/ml



#17 4,4'-DDT


R.T.: 6.181 min
Delta R.T.: -0.010 min
Response: 306587367
Conc: 41.27 ng/ml

Instrument :
ECD_L
ClientSampleId :




#17 4,4'-DDT

R.T.: 7.021 min
Delta R.T.: -0.002 min
Response: 94416320
Conc: 44.28 ng/ml




#18 Endrin aldehyde

R.T.: 6.255 min
Delta R.T.: -0.010 min
Response: 289047653
Conc: 45.39 ng/ml



#18 Endrin aldehyde


R.T.: 6.926 min
Delta R.T.: -0.001 min
Response: 98419501
Conc: 50.44 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.466 min
Delta R.T.: -0.009 min
Response: 338119999
Conc: 44.76 ng/ml

Instrument :
ECD_L
ClientSampleId :




#19 Endosulfan Sulfate

R.T.: 7.148 min
Delta R.T.: -0.001 min
Response: 108804484
Conc: 49.92 ng/ml




#20 Methoxychlor

R.T.: 6.729 min
Delta R.T.: -0.010 min
Response: 161779981
Conc: 42.05 ng/ml



#20 Methoxychlor


R.T.: 7.477 min
Delta R.T.: -0.001 min
Response: 54420567
Conc: 47.79 ng/ml



#21 Endrin ketone


R.T.: 6.949 min
Delta R.T.: -0.010 min
Response: 362153498
Conc: 46.16 ng/ml

Instrument :
ECD_L
ClientSampleId :




#21 Endrin ketone

R.T.: 7.612 min
Delta R.T.: -0.001 min
Response: 114037939
Conc: 47.84 ng/ml




#23 Chlordane-1

R.T.: 0.000 min
Exp R.T. : 4.169 min
Response: 0
Conc: N.D.



#23 Chlordane-1


R.T.: 4.963 min
Delta R.T.: 0.051 min
Response: 184021365
Conc: 1587.97 ng/ml



#24 Chlordane-2


R.T.: 0.000 min
Exp R.T.: 4.663 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#24 Chlordane-2

R.T.: 5.399 min
Delta R.T.: 0.024 min
Response: 168381953
Conc: 1406.11 ng/ml




#25 Chlordane-3

R.T.: 5.211 min
Delta R.T.: -0.008 min
Response: 443501563
Conc: 381.23 ng/ml



#25 Chlordane-3


R.T.: 6.011 min
Delta R.T.: -0.001 min
Response: 150669315
Conc: 369.38 ng/ml



#26 Chlordane-4


R.T.: 5.267 min
Delta R.T.: -0.008 min
Response: 425570510
Conc: 404.77 ng/ml

Instrument :
ECD_L
ClientSampleId :




#26 Chlordane-4

R.T.: 6.085 min
Delta R.T.: -0.004 min
Response: 147965496
Conc: 301.16 ng/ml




#27 Chlordane-5

R.T.: 5.947 min
Delta R.T.: 0.064 min
Response: 356072588
Conc: 2223.60 ng/ml



#27 Chlordane-5


R.T.: 6.926 min
Delta R.T.: 0.044 min
Response: 98419501
Conc: 1098.68 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.991 min
Delta R.T.: -0.012 min
Response: 285995438
Conc: 37.70 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 8.935 min
Delta R.T.: -0.003 min
Response: 109920087
Conc: 44.97 ng/ml