

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080222\
 Data File : PL076846.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Aug 2022 15:30
 Operator : AR\AJ
 Sample : N3996-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 MH-CS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 03 08:45:43 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL080122.M
 Quant Title : GC Extractables
 QLast Update : Mon Aug 01 16:34:20 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.420	2.640	12826844	32940788	13.474	14.272
28) SA Decachlor...	8.911	7.747	11092716	27666553	13.959	14.390

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080222\
Data File : PL076846.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Aug 2022 15:30
Operator : AR\AJ
Sample : N3996-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
MH-CS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 03 08:45:43 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL080122.M
Quant Title : GC Extractables
QLast Update : Mon Aug 01 16:34:20 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.420 min
Delta R.T.: -0.001 min
Response: 12826844
Conc: 13.47 ng/ml

Instrument :
ECD_L
ClientSampleId :
MH-CS




#1 Tetrachloro-m-xylene

R.T.: 2.640 min
Delta R.T.: -0.002 min
Response: 32940788
Conc: 14.27 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.911 min
Delta R.T.: -0.004 min
Response: 11092716
Conc: 13.96 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.747 min
Delta R.T.: -0.004 min
Response: 27666553
Conc: 14.39 ng/ml