

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080224\
 Data File : PL090764.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Aug 2024 17:42
 Operator : AR\AJ
 Sample : P3448-11
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP6

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 08/05/2024
 Supervised By :Ankita Jodhani 08/05/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 03 00:38:23 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071524.M
 Quant Title : GC Extractables
 QLast Update : Mon Jul 15 14:16:48 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.544	2.786	35303521	31394757	16.955	18.654
28) SA Decachlor...	9.050	7.926	21013385	22911395	15.813m	15.713

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080224\
Data File : PL090764.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Aug 2024 17:42
Operator : AR\AJ
Sample : P3448-11
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
TP6


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 08/05/2024
Supervised By : Ankita Jodhani 08/05/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 03 00:38:23 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071524.M
Quant Title : GC Extractables
QLast Update : Mon Jul 15 14:16:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: -0.003 min
Response: 35303521
Conc: 16.95 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP6


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 08/05/2024
Supervised By :Ankita Jodhani 08/05/2024




#1 Tetrachloro-m-xylene

R.T.: 2.786 min
Delta R.T.: -0.003 min
Response: 31394757
Conc: 18.65 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.050 min
Delta R.T.: -0.012 min
Response: 21013385
Conc: 15.81 ng/ml m



#28 Decachlorobiphenyl

R.T.: 7.926 min
Delta R.T.: -0.008 min
Response: 22911395
Conc: 15.71 ng/ml