

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080521\
 Data File : PL068752.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Aug 2021 00:54
 Operator : AR\AJ
 Sample : M3279-02
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 20210569-COM-42

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 06 03:19:54 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL080521.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 06 03:02:43 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.638	5.568	5599078	14257072	13.536	14.042
28)	SA Decachlor...	10.409	11.488	5460190	9128137	9.501	9.408
Target Compounds							
8)	B Heptachlo...	0.000	8.034	0	189881	N.D.	0.164 #
10)	B gamma-Chl...	0.000	8.322f	0	122214	N.D.	0.107 #
15)	B Endosulfa...	0.000	9.216	0	782116	N.D.	0.853 #
23)	Chlordane-1	0.000	6.980f	0	106551	N.D.	2.527 #
25)	Chlordane-3	0.000	8.322f	0	122214	N.D.	0.748 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080521\
Data File : PL068752.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Aug 2021 00:54
Operator : AR\AJ
Sample : M3279-02
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
20210569-COM-42

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 06 03:19:54 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL080521.M
Quant Title : GC Extractables
QLast Update : Fri Aug 06 03:02:43 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 4.638 min
Delta R.T.: -0.002 min
Response: 5599078
Conc: 13.54 ng/ml

Instrument :
ECD_L
ClientSampleId :
20210569-COM-42




#1 Tetrachloro-m-xylene

R.T.: 5.568 min
Delta R.T.: -0.003 min
Response: 14257072
Conc: 14.04 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 7.084 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 8.034 min
Delta R.T.: 0.000 min
Response: 189881
Conc: 0.16 ng/ml



#10 gamma-Chlordane


R.T.: 0.000 min
Exp R.T.: 7.361 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
20210569-COM-42




#10 gamma-Chlordane

R.T.: 8.322 min
Delta R.T.: 0.019 min
Response: 122214
Conc: 0.11 ng/ml




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T.: 8.375 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 9.216 min
Delta R.T.: 0.014 min
Response: 782116
Conc: 0.85 ng/ml



#23 Chlordane-1


R.T.: 0.000 min
Exp R.T.: 5.971 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
20210569-COM-42




#23 Chlordane-1

R.T.: 6.980 min
Delta R.T.: 0.024 min
Response: 106551
Conc: 2.53 ng/ml




#25 Chlordane-3

R.T.: 0.000 min
Exp R.T.: 7.362 min
Response: 0
Conc: N.D.



#25 Chlordane-3


R.T.: 8.322 min
Delta R.T.: 0.018 min
Response: 122214
Conc: 0.75 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.409 min
Delta R.T.: -0.005 min
Response: 5460190
Conc: 9.50 ng/ml

Instrument :
ECD_L
ClientSampleId :
20210569-COM-42



#28 Decachlorobiphenyl

R.T.: 11.488 min
Delta R.T.: -0.006 min
Response: 9128137
Conc: 9.41 ng/ml