

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080522\
 Data File : PL076933.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Aug 2022 16:49
 Operator : AR\AJ
 Sample : PB146783BL
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB146783BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 05 22:28:13 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL080122.M
 Quant Title : GC Extractables
 QLast Update : Mon Aug 01 16:34:20 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.419	2.638	17401912	45721569	18.279	19.809
28)	SA Decachlor...	8.906	7.742	16252127	41196623	20.452	21.428

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080522\
Data File : PL076933.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Aug 2022 16:49
Operator : AR\AJ
Sample : PB146783BL
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB146783BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 05 22:28:13 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL080122.M
Quant Title : GC Extractables
QLast Update : Mon Aug 01 16:34:20 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.419 min
Delta R.T.: -0.002 min
Response: 17401912
Conc: 18.28 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB146783BL




#1 Tetrachloro-m-xylene

R.T.: 2.638 min
Delta R.T.: -0.004 min
Response: 45721569
Conc: 19.81 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.906 min
Delta R.T.: -0.009 min
Response: 16252127
Conc: 20.45 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.742 min
Delta R.T.: -0.009 min
Response: 41196623
Conc: 21.43 ng/ml