

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080618\
 Data File : PL038950.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Aug 2018 12:53
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 8/7/2018 2:36:23 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 07 02:04:43 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
 Quant Title : GC Extractables
 QLast Update : Sat Jul 14 00:12:12 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.323	3.927	314.1E6	123.7E6	42.177	47.109
28)	SA Decachlor...	7.987	8.932	315.7E6	110.9E6	41.617m	45.358
Target Compounds							
2)	A alpha-BHC	3.752	4.306	520.9E6	195.7E6	44.083	50.558
3)	MA gamma-BHC...	4.027	4.585	476.8E6	175.4E6	45.478m	48.549
4)	MA Heptachlor	4.311	5.095	482.0E6	173.1E6	45.308	54.152
5)	MB Aldrin	4.550	5.397	472.2E6	162.4E6	45.110	52.607
6)	B beta-BHC	4.283	4.748	211.9E6	80469058	44.899	53.049
7)	B delta-BHC	4.475	4.962	488.8E6	173.1E6	45.031m	52.198
8)	B Heptachlo...	4.985	5.777	438.8E6	149.2E6	46.537m	52.505
9)	A Endosulfan I	5.318	6.130	411.8E6	129.8E6	47.608	50.477
10)	B gamma-Chl...	5.209	6.010	458.9E6	146.7E6	46.426	50.599
11)	B alpha-Chl...	5.265	6.083	443.4E6	147.2E6	46.845	51.370
12)	B 4,4'-DDE	5.432	6.239	435.4E6	138.7E6	45.408m	51.852
13)	MA Dieldrin	5.556	6.383	446.1E6	136.3E6	46.420	50.095
14)	MA Endrin	5.808	6.598	406.7E6	119.2E6	46.806m	49.903
15)	B Endosulfa...	6.083	6.800	386.5E6	112.9E6	48.739m	47.027m
16)	A 4,4'-DDD	5.944	6.721	368.9E6	99338625	50.289	46.411
17)	MA 4,4'-DDT	6.178	7.020	353.2E6	103.3E6	47.552	48.464
18)	B Endrin al...	6.252	6.924	302.5E6	97355250	47.504	49.891
19)	B Endosulfa...	6.463	7.145	368.6E6	108.1E6	48.794	49.608m
20)	A Methoxychlor	6.726	7.476	188.3E6	58653965	48.931	51.511
21)	B Endrin ke...	6.944	7.611	400.4E6	113.0E6	51.033m	47.400

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080618\
Data File : PL038950.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Aug 2018 12:53
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
8/7/2018 2:36:23 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 07 02:04:43 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
Quant Title : GC Extractables
QLast Update : Sat Jul 14 00:12:12 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

