

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080619\
 Data File : PL051093.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Aug 2019 12:40
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 8/7/2019 5:44:12 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 07 00:36:34 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 19 01:19:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.319	3.928	585.4E6	149.8E6	58.261	52.494
28)	SA Decachlor...	7.980	8.958	576.5E6	148.2E6	56.939	52.421
Target Compounds							
2)	A alpha-BHC	3.746	4.316	989.4E6	224.7E6	59.044	54.458
3)	MA gamma-BHC...	4.022	4.599	922.4E6	212.4E6	59.968	54.848
4)	MA Heptachlor	4.306	5.101	968.3E6	223.6E6	59.015	53.928
5)	MB Aldrin	4.544	5.403	927.9E6	208.1E6	59.899	55.229
6)	B beta-BHC	4.277	4.776	383.4E6	97618530	58.489	56.464
7)	B delta-BHC	4.470	4.989	974.7E6	214.7E6	59.401	55.520
8)	B Heptachlor...	4.980	5.790	864.7E6	200.5E6	58.529	52.432
9)	A Endosulfan I	5.311	6.144	812.9E6	188.2E6	55.895	54.736
10)	B gamma-Chl...	5.202	6.024	906.6E6	204.1E6	57.442	55.229
11)	B alpha-Chl...	5.258	6.095	873.8E6	202.0E6	57.275	55.102
12)	B 4,4'-DDE	5.426	6.255	794.2E6	201.1E6	52.994	55.807
13)	MA Dieldrin	5.549	6.401	887.3E6	198.4E6	57.257	56.038
14)	MA Endrin	5.802	6.618	797.9E6	167.8E6	54.206	54.089
15)	B Endosulfa...	6.077	6.830	765.4E6	174.7E6	54.692	54.943
16)	A 4,4'-DDD	5.936	6.746	710.7E6	154.6E6	59.305	58.060
17)	MA 4,4'-DDT	6.170	7.043	657.7E6	146.7E6	53.102	54.328
18)	B Endrin al...	6.244	6.953	621.3E6	147.6E6	53.855	54.304m
19)	B Endosulfa...	6.455	7.177	686.1E6	162.7E6	53.860	56.348
20)	A Methoxychlor	6.715	7.504	326.4E6	78724947	52.311	53.654
21)	B Endrin ke...	6.936	7.646	760.8E6	167.4E6	55.621	57.914
22)	Mirex	7.109	8.094	530.0E6	124.6E6	54.059	54.261

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080619\
Data File : PL051093.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Aug 2019 12:40
Operator : SM\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
8/7/2019 5:44:12 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 07 00:36:34 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
Quant Title : GC Extractables
QLast Update : Fri Jul 19 01:19:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

