

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080619\
 Data File : PL051115.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Aug 2019 18:19
 Operator : SM\AJ
 Sample : K4162-04MS
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 RBR-200050MS

Manual Integrations
 APPROVED

Ankita
 8/7/2019 5:45:01 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 07 00:44:08 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 19 01:19:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.318	3.928	239.3E6	68315256	23.816	23.943
28)	SA Decachlor...	7.979	8.958	216.5E6	59470754	21.380	21.030
Target Compounds							
2)	A alpha-BHC	3.745	4.317	1146.9E6	303.3E6	68.439	73.512
3)	MA gamma-BHC...	4.021	4.600	1078.5E6	281.1E6	70.119	72.570
4)	MA Heptachlor	4.305	5.101	1131.0E6	286.0E6	68.928	68.965
5)	MB Aldrin	4.543	5.403	1046.9E6	262.9E6	67.578	69.754
6)	B beta-BHC	4.276	4.776	462.4E6	126.9E6	70.536	73.425
7)	B delta-BHC	4.468	4.990	1056.4E6	238.6E6	64.381	61.683
8)	B Heptachlor...	4.979	5.790	1005.4E6	251.5E6	68.057	65.756
9)	A Endosulfan I	5.309	6.144	948.3E6	239.5E6	65.204	69.633
10)	B gamma-Chl...	5.201	6.024	1063.2E6	251.9E6	67.361	68.165
11)	B alpha-Chl...	5.257	6.095	1013.2E6	252.8E6	66.410	68.975
12)	B 4,4'-DDE	5.424	6.256	966.8E6	246.3E6	64.513	68.352
13)	MA Dieldrin	5.548	6.401	1040.3E6	253.0E6	67.130	71.457
14)	MA Endrin	5.801	6.618	983.3E6	217.7E6	66.798	70.179
15)	B Endosulfa...	6.076	6.831	877.3E6	220.3E6	62.688	69.283
16)	A 4,4'-DDD	5.935	6.746	776.8E6	188.1E6	64.819	70.643
17)	MA 4,4'-DDT	6.169	7.043	770.4E6	183.3E6	62.199	67.885
18)	B Endrin al...	6.242	6.953	660.9E6	166.5E6	57.285	61.260m
19)	B Endosulfa...	6.454	7.178	758.2E6	192.8E6	59.515	66.767
20)	A Methoxychlor	6.714	7.504	377.8E6	97455289	60.558	66.420
21)	B Endrin ke...	6.935	7.646	856.6E6	209.3E6	62.626	72.375
22)	Mirex	7.108	8.094	527.0E6	135.9E6	53.756	59.183

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080619\
Data File : PL051115.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Aug 2019 18:19
Operator : SM\AJ
Sample : K4162-04MS
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
RBR-200050MS

Manual Integrations
APPROVED

Ankita
8/7/2019 5:45:01 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 07 00:44:08 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
Quant Title : GC Extractables
QLast Update : Fri Jul 19 01:19:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

