

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080619\
 Data File : PL051142.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Aug 2019 01:08
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 8/7/2019 5:46:04 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 07 04:41:01 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 19 01:19:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.318	3.929	527.7E6	142.5E6	52.520	49.954
28)	SA Decachlor...	7.979	8.958	567.2E6	148.0E6	56.014	52.349
Target Compounds							
2)	A alpha-BHC	3.746	4.317	885.6E6	215.8E6	52.845	52.312
3)	MA gamma-BHC...	4.021	4.600	828.0E6	203.0E6	53.829	52.420
4)	MA Heptachlor	4.305	5.101	862.1E6	210.9E6	52.545	50.859
5)	MB Aldrin	4.543	5.403	840.0E6	200.7E6	54.220	53.264
6)	B beta-BHC	4.276	4.777	347.1E6	93846958	52.941	54.283
7)	B delta-BHC	4.469	4.990	879.6E6	205.1E6	53.607	53.038
8)	B Heptachlor...	4.979	5.790	781.3E6	193.3E6	52.885	50.532
9)	A Endosulfan I	5.309	6.145	743.0E6	182.1E6	51.085	52.963
10)	B gamma-Chl...	5.201	6.023	826.6E6	197.4E6	52.369	53.422
11)	B alpha-Chl...	5.257	6.095	794.6E6	194.6E6	52.085	53.085
12)	B 4,4'-DDE	5.425	6.255	739.8E6	193.4E6	49.369	53.664
13)	MA Dieldrin	5.548	6.401	822.5E6	192.3E6	53.077	54.314
14)	MA Endrin	5.800	6.618	721.6E6	157.2E6	49.021	50.675
15)	B Endosulfa...	6.075	6.830	718.2E6	169.6E6	51.321	53.336
16)	A 4,4'-DDD	5.935	6.746	666.7E6	152.4E6	55.633	57.235
17)	MA 4,4'-DDT	6.168	7.043	597.6E6	137.7E6	48.250	51.026
18)	B Endrin al...	6.242	6.953	587.2E6	143.3E6	50.896	52.736m
19)	B Endosulfa...	6.453	7.177	655.4E6	157.5E6	51.445	54.533
20)	A Methoxychlor	6.713	7.504	303.7E6	75017120	48.681	51.127
21)	B Endrin ke...	6.935	7.646	740.9E6	169.4E6	54.172	58.601
22)	Mirex	7.108	8.094	499.8E6	122.1E6	50.980	53.141

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080619\
Data File : PL051142.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Aug 2019 01:08
Operator : SM\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
8/7/2019 5:46:04 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 07 04:41:01 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
Quant Title : GC Extractables
QLast Update : Fri Jul 19 01:19:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

