

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : PL080824CLP.M
 Title : GC Extractables
 Last Update : Sat Aug 10 06:55:30 2024
 Response Via : Initial Calibration

Calibration Files

3 =PL090865.D 1 =PL090863.D 2 =PL090864.D
 4 =PL090866.D 5 =PL090867.D

	Compound	3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.167	2.259	2.191	2.037	2.039	2.138	E6	4.58
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1	7.478	8.490	7.908	6.919	6.688	7.497	E3	9.77
23)	Toxaphene-2	1.586	2.005	2.180	1.805	1.803	1.876	E4	12.03
24)	Toxaphene-3	4.948	5.410	5.082	4.780	4.945	5.033	E4	4.70
25)	Toxaphene-4	3.732	4.100	3.802	3.560	3.764	3.792	E4	5.16
26)	Toxaphene-5	2.686	2.797	2.685	2.587	2.722	2.695	E4	2.81
27) SA	Decachlorobiphenyl	8.948	9.443	9.282	8.133	8.397	8.841	E5	6.37

Signal #2 Calibration Files

3 =PL090865.D 1 =PL090863.D 2 =PL090864.D
 4 =PL090866.D 5 =PL090867.D

	Compound	3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.829	1.812	1.800	1.819	1.905	1.833	E6	2.28
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1	1.108	1.137	1.127	1.107	1.170	1.130	E4	2.30

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
Method File : PL080824CLP.M
Title : GC Extractables
Last Update : Sat Aug 10 06:55:30 2024
Response Via : Initial Calibration

Calibration Files

3	=PL090865.D	1	=PL090863.D	2	=PL090864.D
4	=PL090866.D	5	=PL090867.D		

	Compound	3	1	2	4	5	Avg	%RSD
23)	Toxaphene-2	1.003	0.881	0.976	1.086	1.181	1.025	E4 11.10
24)	Toxaphene-3	3.876	3.299	3.393	4.088	4.575	3.846	E4 13.60
25)	Toxaphene-4	5.296	4.692	4.944	5.532	5.943	5.281	E4 9.29
26)	Toxaphene-5	3.517	3.207	3.152	3.570	3.802	3.450	E4 7.82
27) SA	Decachlorobiphenyl	8.825	8.977	8.678	8.472	9.094	8.809	E5 2.79

(#) = Out of Range