

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080824\
 Data File : PL090874.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Aug 2024 11:53
 Operator : AR\AJ
 Sample : INDA401
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 INDA4311

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 08 14:01:16 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL080824CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Aug 08 13:53:03 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.544	2.786	78836713	70607704	40.304	40.317
2) SA Decachlor...	9.053	7.925	116.2E6	133.1E6	80.495	81.047
Target Compounds						
2) A alpha-BHC	3.999	3.289	115.3E6	105.3E6	40.125	40.509
3) MA gamma-BHC...	4.331	3.619	108.8E6	98404721	40.179	40.497
4) MA Heptachlor	4.920	3.959	100.1E6	94971299	40.072	40.394
9) A Endosulfan I	6.072	5.110	82824706	67031628	40.074	38.467
13) MA Dieldrin	6.347	5.375	172.3E6	175.7E6	80.158	81.314
14) MA Endrin	6.576	5.650	139.1E6	155.3E6	80.463	81.649
16) A 4,4'-DDD	6.711	5.798	129.6E6	135.0E6	80.374	81.140
17) MA 4,4'-DDT	7.024	6.049	125.1E6	136.1E6	80.265	81.317
20) A Methoxychlor	7.499	6.623	317.5E6	349.3E6	404.280	407.685

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080824\
Data File : PL090874.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Aug 2024 11:53
Operator : AR\AJ
Sample : INDA401
Misc :
ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 08 14:01:16 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL080824CLP.M
Quant Title : GC Extractables
QLast Update : Thu Aug 08 13:53:03 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

