

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080918\
 Data File : PL039075.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Aug 2018 17:15
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 8/10/2018 10:46:36 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 01:12:41 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
 Quant Title : GC Extractables
 QLast Update : Sat Jul 14 00:12:12 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.323	3.927	326.4E6	132.7E6	43.832	50.539
28)	SA Decachlor...	7.989	8.934	304.1E6	112.9E6	40.082m	46.167
Target Compounds							
2)	A alpha-BHC	3.752	4.306	549.9E6	210.7E6	46.538	54.436
3)	MA gamma-BHC...	4.027	4.586	500.3E6	189.6E6	47.718m	52.497
4)	MA Heptachlor	4.311	5.096	497.8E6	181.7E6	46.801	56.839
5)	MB Aldrin	4.549	5.398	491.1E6	171.0E6	46.916	55.403
6)	B beta-BHC	4.283	4.749	221.8E6	85865551	46.997	56.606
7)	B delta-BHC	4.476	4.963	520.3E6	189.3E6	47.935	57.074
8)	B Heptachlo...	4.986	5.778	455.6E6	156.8E6	48.326m	55.159
9)	A Endosulfan I	5.318	6.130	419.8E6	135.8E6	48.533	52.812
10)	B gamma-Chl...	5.209	6.010	474.8E6	152.6E6	48.038	52.623
11)	B alpha-Chl...	5.266	6.084	454.1E6	150.9E6	47.971	52.639
12)	B 4,4'-DDE	5.433	6.240	447.4E6	144.8E6	46.663m	54.156
13)	MA Dieldrin	5.557	6.383	455.6E6	140.1E6	47.412	51.468m
14)	MA Endrin	5.810	6.599	403.1E6	121.9E6	46.394	51.029
15)	B Endosulfa...	6.084	6.803	391.8E6	121.0E6	49.410m	50.381
16)	A 4,4'-DDD	5.944	6.722	366.3E6	108.4E6	49.936m	50.665
17)	MA 4,4'-DDT	6.179	7.021	338.7E6	101.9E6	45.595	47.810
18)	B Endrin al...	6.253	6.925	304.7E6	101.7E6	47.847	52.131
19)	B Endosulfa...	6.464	7.148	357.7E6	111.5E6	47.356	51.172
20)	A Methoxychlor	6.727	7.477	177.9E6	57401297	46.240	50.411
21)	B Endrin ke...	6.946	7.610	396.8E6	118.0E6	50.568m	49.527m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL080918\
Data File : PL039075.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Aug 2018 17:15
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
8/10/2018 10:46:36 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 01:12:41 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL071318.M
Quant Title : GC Extractables
QLast Update : Sat Jul 14 00:12:12 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

