

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081219\
 Data File : PL051297.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Aug 2019 12:41
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 8/14/2019 9:32:40 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 13 00:42:50 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 19 01:19:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.316	3.927	557.5E6	142.0E6	55.489	49.768
28)	SA Decachlor...	7.974	8.954	573.8E6	133.8E6	56.673m	47.300
Target Compounds							
2)	A alpha-BHC	3.743	4.315	930.1E6	214.6E6	55.500	52.001
3)	MA gamma-BHC...	4.018	4.598	873.6E6	202.3E6	56.795	52.237
4)	MA Heptachlor	4.302	5.099	922.0E6	212.7E6	56.193	51.305
5)	MB Aldrin	4.539	5.401	869.0E6	199.1E6	56.097	52.822
6)	B beta-BHC	4.273	4.774	364.9E6	93197529	55.667	53.907
7)	B delta-BHC	4.466	4.988	924.2E6	203.6E6	56.323	52.644
8)	B Heptachlo...	4.976	5.788	804.2E6	188.3E6	54.433	49.225
9)	A Endosulfan I	5.306	6.142	772.9E6	178.0E6	53.140	51.751
10)	B gamma-Chl...	5.198	6.021	856.5E6	193.7E6	54.269	52.429
11)	B alpha-Chl...	5.254	6.092	825.8E6	191.7E6	54.129	52.317
12)	B 4,4'-DDE	5.422	6.253	802.0E6	192.0E6	53.519	53.285
13)	MA Dieldrin	5.545	6.398	840.9E6	188.1E6	54.266	53.110
14)	MA Endrin	5.797	6.616	793.8E6	160.7E6	53.924	51.809
15)	B Endosulfa...	6.073	6.828	722.7E6	164.2E6	51.639	51.660
16)	A 4,4'-DDD	5.932	6.742	663.4E6	147.5E6	55.358	55.380m
17)	MA 4,4'-DDT	6.165	7.040	667.5E6	147.9E6	53.896	54.772
18)	B Endrin al...	6.239	6.950	585.3E6	137.2E6	50.736	50.491m
19)	B Endosulfa...	6.450	7.175	647.6E6	151.9E6	50.832	52.586
20)	A Methoxychlor	6.711	7.501	323.6E6	82343085	51.865	56.120
21)	B Endrin ke...	6.932	7.643	656.5E6	155.6E6	47.999	53.800
22)	Mirex	7.105	8.091	543.0E6	113.7E6	55.379	49.505

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081219\
Data File : PL051297.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2019 12:41
Operator : SM\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
8/14/2019 9:32:40 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 00:42:50 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
Quant Title : GC Extractables
QLast Update : Fri Jul 19 01:19:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

