

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX081324.M
 Title : GC Extractables
 Last Update : Wed Aug 14 06:43:54 2024
 Response Via : Initial Calibration

Calibration Files

500 =PL091019.D 1000 =PL091017.D 250 =PL091020.D
 750 =PL091018.D 100 =PL091021.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.007	1.968	2.105	1.949	2.450	2.096	E6	9.87
2)	Toxaphene-1	1.039	0.969	1.062	0.985	1.151	1.041	E4	6.93
3)	Toxaphene-2	1.269	1.165	1.402	1.143	1.483	1.293	E4	11.44
4)	Toxaphene-3	6.570	6.144	7.022	6.211	8.245	6.839	E4	12.58
5)	Toxaphene-4	4.884	4.608	5.322	4.613	6.305	5.146	E4	13.79
6)	Toxaphene-5	3.750	3.513	4.009	3.558	4.662	3.899	E4	12.05
7) SA	Decachlorobiphenyl	1.432	1.338	1.551	1.357	1.844	1.504	E6	13.81

Signal #2 Calibration Files

500 =PL091019.D 1000 =PL091017.D 250 =PL091020.D
 750 =PL091018.D 100 =PL091021.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	1.871	1.923	1.888	1.858	2.059	1.920	E6	4.23
2)	Toxaphene-1	1.383	1.342	1.426	1.320	1.569	1.408	E4	7.02
3)	Toxaphene-2	1.400	1.431	1.385	1.352	1.441	1.402	E4	2.56
4)	Toxaphene-3	4.560	4.869	4.441	4.607	4.651	4.626	E4	3.39
5)	Toxaphene-4	6.480	6.744	6.258	6.508	6.746	6.547	E4	3.13
6)	Toxaphene-5	2.050	2.173	1.947	2.077	1.979	2.045	E4	4.33
7) SA	Decachlorobiphenyl	1.796	1.748	1.857	1.734	2.142	1.856	E6	9.02

(#) = Out of Range