

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081419\
 Data File : PL051484.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Aug 2019 18:06
 Operator : SM\AJ
 Sample : PB122247BS
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB122247BS

Manual Integrations
 APPROVED

Ankita
 8/16/2019 2:42:28 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 15 09:10:55 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 19 01:19:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.316	3.927	200.1E6	57331942	19.918	20.094
28)	SA Decachlor...	7.975	8.955	227.3E6	58095188	22.452	20.543
Target Compounds							
2)	A alpha-BHC	3.743	4.315	843.7E6	208.3E6	50.345	50.488
3)	MA gamma-BHC...	4.018	4.598	794.0E6	200.5E6	51.617	51.757
4)	MA Heptachlor	4.302	5.099	808.3E6	217.6E6	49.261	52.480
5)	MB Aldrin	4.540	5.401	823.6E6	203.4E6	53.166	53.965
6)	B beta-BHC	4.273	4.775	322.1E6	93634892	49.136	54.160
7)	B delta-BHC	4.466	4.988	698.0E6	168.6E6	42.541	43.593
8)	B Heptachlor...	4.976	5.788	753.6E6	194.3E6	51.012	50.797
9)	A Endosulfan I	5.305	6.143	736.7E6	183.2E6	50.653	53.263
10)	B gamma-Chl...	5.198	6.022	805.7E6	197.8E6	51.051	53.537
11)	B alpha-Chl...	5.253	6.093	770.9E6	196.3E6	50.530	53.565
12)	B 4,4'-DDE	5.421	6.254	776.2E6	197.9E6	51.796	54.922
13)	MA Dieldrin	5.545	6.399	789.8E6	194.3E6	50.966	54.860
14)	MA Endrin	5.797	6.616	688.9E6	151.8E6	46.802	48.949
15)	B Endosulfa...	6.072	6.828	680.6E6	169.7E6	48.633	53.369
16)	A 4,4'-DDD	5.931	6.743	638.7E6	158.6E6	53.299	59.546m
17)	MA 4,4'-DDT	6.165	7.041	581.8E6	144.7E6	46.977	53.616
18)	B Endrin al...	6.239	6.951	565.9E6	145.6E6	49.058	53.561m
19)	B Endosulfa...	6.451	7.176	653.9E6	166.1E6	51.332	57.515
20)	A Methoxychlor	6.710	7.502	295.6E6	78390122	47.376	53.426
21)	B Endrin ke...	6.932	7.645	745.7E6	180.0E6	54.520	62.241
22)	Mirex	7.105	8.092	499.8E6	126.5E6	50.975	55.061

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081419\
Data File : PL051484.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Aug 2019 18:06
Operator : SM\AJ
Sample : PB122247BS
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PB122247BS

Manual Integrations
APPROVED

Ankita
8/16/2019 2:42:28 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 15 09:10:55 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
Quant Title : GC Extractables
QLast Update : Fri Jul 19 01:19:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

