

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081419\
 Data File : PL051490.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Aug 2019 19:31
 Operator : SM\AJ
 Sample : K4340-03MSD
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 LS-SAMPLE-2MSD

Manual Integrations
 APPROVED

Ankita
 8/16/2019 2:42:42 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 15 09:12:37 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 19 01:19:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.316	3.927	210.4E6	58509777	20.942	20.507
28)	SA Decachlor...	7.976	8.955	153.4E6	39824970	15.152	14.083
Target Compounds							
2)	A alpha-BHC	3.743	4.315	1021.2E6	250.9E6	60.939	60.820
3)	MA gamma-BHC...	4.018	4.598	959.1E6	239.5E6	62.352	61.830
4)	MA Heptachlor	4.302	5.099	1002.0E6	244.5E6	61.071	58.967
5)	MB Aldrin	4.540	5.401	957.0E6	234.3E6	61.779	62.168
6)	B beta-BHC	4.273	4.775	406.6E6	109.5E6	62.017	63.360
7)	B delta-BHC	4.466	4.988	825.7E6	185.2E6	50.323	47.889
8)	B Heptachlor...	4.976	5.788	880.4E6	216.3E6	59.590	56.552
9)	A Endosulfan I	5.306	6.143	839.4E6	206.9E6	57.712	60.151
10)	B gamma-Chl...	5.198	6.022	909.6E6	216.2E6	57.633	58.500
11)	B alpha-Chl...	5.254	6.093	882.9E6	218.1E6	57.871	59.505
12)	B 4,4'-DDE	5.421	6.253	865.7E6	218.2E6	57.768	60.560
13)	MA Dieldrin	5.544	6.398	907.9E6	218.6E6	58.586	61.742
14)	MA Endrin	5.797	6.616	843.0E6	183.6E6	57.269	59.189
15)	B Endosulfa...	6.072	6.828	762.2E6	190.3E6	54.460	59.870
16)	A 4,4'-DDD	5.931	6.743	701.3E6	167.1E6	58.524	62.764m
17)	MA 4,4'-DDT	6.165	7.041	683.4E6	164.8E6	55.177	61.034
18)	B Endrin al...	6.239	6.951	629.1E6	157.2E6	54.531	57.860m
19)	B Endosulfa...	6.451	7.176	732.8E6	183.6E6	57.523	63.575
20)	A Methoxychlor	6.710	7.502	356.2E6	91825803	57.097	62.583
21)	B Endrin ke...	6.932	7.645	839.6E6	194.1E6	61.388	67.140
22)	Mirex	7.105	8.092	554.9E6	134.8E6	56.602	58.671

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081419\
Data File : PL051490.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Aug 2019 19:31
Operator : SM\AJ
Sample : K4340-03MSD
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
LS-SAMPLE-2MSD

Manual Integrations
APPROVED

Ankita
8/16/2019 2:42:42 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 15 09:12:37 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
Quant Title : GC Extractables
QLast Update : Fri Jul 19 01:19:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

