

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081419\
 Data File : PL051496.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Aug 2019 20:56
 Operator : SM\AJ
 Sample : K4089-02MS
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 OK-03-081219MS

Manual Integrations
 APPROVED

Ankita
 8/16/2019 2:42:46 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 15 09:14:14 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 19 01:19:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.316	3.928	219.5E6	66039186	21.848	23.146
28)	SA Decachlor...	7.975	8.955	248.9E6	65334473	24.581	23.103
Target Compounds							
2)	A alpha-BHC	3.743	4.315	1016.8E6	266.1E6	60.674	64.485
3)	MA gamma-BHC...	4.018	4.598	961.0E6	253.0E6	62.474	65.319
4)	MA Heptachlor	4.301	5.099	1011.9E6	261.3E6	61.672	63.009
5)	MB Aldrin	4.539	5.401	1001.4E6	258.2E6	64.645	68.530
6)	B beta-BHC	4.273	4.775	399.4E6	113.6E6	60.915	65.679
7)	B delta-BHC	4.465	4.988	853.4E6	212.1E6	52.008	54.831
8)	B Heptachlor...	4.975	5.788	918.1E6	244.1E6	62.147	63.810
9)	A Endosulfan I	5.306	6.143	901.0E6	231.0E6	61.948	67.170
10)	B gamma-Chl...	5.198	6.022	984.0E6	249.9E6	62.344	67.623
11)	B alpha-Chl...	5.253	6.093	947.7E6	245.6E6	62.118	67.017
12)	B 4,4'-DDE	5.420	6.253	941.4E6	248.0E6	62.822	68.830
13)	MA Dieldrin	5.544	6.399	975.9E6	246.5E6	62.978	69.603
14)	MA Endrin	5.796	6.616	879.7E6	203.9E6	59.757	65.752
15)	B Endosulfa...	6.072	6.828	846.0E6	212.6E6	60.450	66.873
16)	A 4,4'-DDD	5.931	6.742	774.5E6	195.9E6	64.627	73.585m
17)	MA 4,4'-DDT	6.165	7.041	717.9E6	179.1E6	57.960	66.330
18)	B Endrin al...	6.238	6.951	684.2E6	176.5E6	59.307	64.943m
19)	B Endosulfa...	6.450	7.175	802.1E6	207.3E6	62.963	71.773
20)	A Methoxychlor	6.709	7.501	352.0E6	95431508	56.426	65.040
21)	B Endrin ke...	6.931	7.643	885.1E6	217.9E6	64.714	75.365
22)	Mirex	7.104	8.091	618.2E6	158.0E6	63.058	68.786

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081419\
Data File : PL051496.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Aug 2019 20:56
Operator : SM\AJ
Sample : K4089-02MS
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
OK-03-081219MS

Manual Integrations
APPROVED

Ankita
8/16/2019 2:42:46 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 15 09:14:14 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
Quant Title : GC Extractables
QLast Update : Fri Jul 19 01:19:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

