

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081518\
 Data File : PL039187.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Aug 2018 07:33
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 8/16/2018 12:35:41 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 09:52:03 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081318.M
 Quant Title : GC Extractables
 QLast Update : Wed Aug 15 15:04:22 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.305	3.907	356.4E6	144.2E6	50.978	53.771
28)	SA Decachlor...	7.962	8.894	359.5E6	123.3E6	51.184	53.124
Target Compounds							
2)	A alpha-BHC	3.733	4.285	603.5E6	229.0E6	51.056	54.370
3)	MA gamma-BHC...	4.009	4.563	544.9E6	207.9E6	51.378	51.644
4)	MA Heptachlor	4.291	5.070	536.2E6	198.5E6	50.932	55.523
5)	MB Aldrin	4.528	5.371	520.8E6	189.7E6	52.000	53.586
6)	B beta-BHC	4.263	4.725	238.2E6	93544080	51.279	53.407
7)	B delta-BHC	4.456	4.938	564.5E6	207.6E6	52.008	52.129
8)	B Heptachlo...	4.964	5.750	487.2E6	177.7E6	51.020	52.851
9)	A Endosulfan I	5.293	6.101	451.7E6	152.2E6	51.587	51.875
10)	B gamma-Chl...	5.185	5.982	508.9E6	170.5E6	51.523	52.691
11)	B alpha-Chl...	5.241	6.055	490.1E6	169.1E6	51.012	52.260
12)	B 4,4'-DDE	5.411	6.211	494.6E6	161.6E6	51.651	52.262
13)	MA Dieldrin	5.533	6.354	504.4E6	158.4E6	51.722	52.273
14)	MA Endrin	5.785	6.568	457.6E6	131.4E6	51.124	51.119
15)	B Endosulfa...	6.060	6.771	446.2E6	132.6E6	51.314	51.614
16)	A 4,4'-DDD	5.921	6.692	403.2E6	119.1E6	50.590	53.169
17)	MA 4,4'-DDT	6.154	6.990	392.5E6	115.8E6	51.783	49.211
18)	B Endrin al...	6.228	6.894	340.8E6	111.9E6	50.292	51.518
19)	B Endosulfa...	6.438	7.116	413.0E6	122.7E6	50.757	51.173
20)	A Methoxychlor	6.702	7.445	207.8E6	63608833	50.790	51.358
21)	B Endrin ke...	6.920	7.579	428.1E6	128.6E6	49.734	52.049

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081518\
Data File : PL039187.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Aug 2018 07:33
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
8/16/2018 12:35:41 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 09:52:03 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081318.M
Quant Title : GC Extractables
QLast Update : Wed Aug 15 15:04:22 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

