

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081519\
 Data File : PL051518.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2019 15:02
 Operator : SG\AJ
 Sample : PB122286BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB122286BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 00:50:15 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 19 01:19:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.324	3.927	196.2E6	54430762	19.529	19.077
28)	SA Decachlor...	7.985	8.958	241.7E6	61856773	23.872	21.874
Target Compounds							
11)	B alpha-Chl...	5.263	0.000	212.6E6	0	13.936	N.D. #
13)	MA Dieldrin	5.541	0.000	3940194	0	0.254	N.D. #
17)	MA 4,4'-DDT	6.178	0.000	46442910	0	3.750	N.D. #
18)	B Endrin al...	6.257	0.000	16498050	0	1.430	N.D. #
21)	B Endrin ke...	6.927	0.000	26211762	0	1.916	N.D. #
26)	Chlordane-4	5.263	0.000	212.6E6	0	161.196	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL051519\
Data File : PL051518.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2019 15:02
Operator : SG\AJ
Sample : PB122286BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleID :
PB122286BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 00:50:15 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071719.M
Quant Title : GC Extractables
QLast Update : Fri Jul 19 01:19:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.324 min
Delta R.T.: 0.006 min
Response: 196228297
Conc: 19.53 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB122286BL




#1 Tetrachloro-m-xylene

R.T.: 3.927 min
Delta R.T.: -0.003 min
Response: 54430762
Conc: 19.08 ng/ml




#11 alpha-Chlordane

R.T.: 5.263 min
Delta R.T.: 0.006 min
Response: 212613996
Conc: 13.94 ng/ml



#11 alpha-Chlordane


R.T.: 0.000 min
Exp R.T.: 6.098 min
Response: 0
Conc: N.D.



#13 Dieldrin


R.T.: 5.541 min
Delta R.T.: -0.007 min
Response: 3940194
Conc: 0.25 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB122286BL




#13 Dieldrin

R.T.: 0.000 min
Exp R.T. : 6.404 min
Response: 0
Conc: N.D.




#17 4,4'-DDT

R.T.: 6.178 min
Delta R.T.: 0.008 min
Response: 46442910
Conc: 3.75 ng/ml



#17 4,4'-DDT


R.T.: 0.000 min
Exp R.T. : 7.046 min
Response: 0
Conc: N.D.



#18 Endrin aldehyde


R.T.: 6.257 min
Delta R.T.: 0.015 min
Response: 16498050
Conc: 1.43 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB122286BL




#18 Endrin aldehyde

R.T.: 0.000 min
Exp R.T. : 6.957 min
Response: 0
Conc: N.D.




#21 Endrin ketone

R.T.: 6.927 min
Delta R.T.: -0.009 min
Response: 26211762
Conc: 1.92 ng/ml



#21 Endrin ketone


R.T.: 0.000 min
Exp R.T. : 7.648 min
Response: 0
Conc: N.D.



#26 Chlordane-4


R.T.: 5.263 min
Delta R.T.: 0.006 min
Response: 212613996
Conc: 161.20 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB122286BL




#26 Chlordane-4

R.T.: 0.000 min
Exp R.T.: 6.100 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 7.985 min
Delta R.T.: 0.005 min
Response: 241710426
Conc: 23.87 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.958 min
Delta R.T.: -0.004 min
Response: 61856773
Conc: 21.87 ng/ml