

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081521\
 Data File : PL068999.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2021 14:24
 Operator : AR\AJ
 Sample : M3404-04
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP2-1-6FT

Manual Integrations
 APPROVED

mohammad
 8/16/2021 1:40:32 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 15 23:18:24 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL080521.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 06 03:02:43 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.636	5.568	6827640	17239869	16.506	16.979
28)	SA Decachlor...	10.404	11.488	10263191	16849536	17.859	17.366
Target Compounds							
12)	B 4,4'-DDE	7.637	8.564	821182	1191045	1.700	1.127 #
16)	A 4,4'-DDD	8.222	9.104	180954	493713	0.422m	0.578m#
17)	MA 4,4'-DDT	8.483	9.423	841941	921712	1.734	0.974 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081521\
Data File : PL068999.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2021 14:24
Operator : AR\AJ
Sample : M3404-04
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampled :
TP2-1-6FT


Manual Integrations
APPROVED

mohammad
8/16/2021 1:40:32 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 15 23:18:24 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL080521.M
Quant Title : GC Extractables
QLast Update : Fri Aug 06 03:02:43 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 4.636 min
Delta R.T.: -0.004 min
Response: 6827640
Conc: 16.51 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP2-1-6FT


Manual Integrations
APPROVED

mohammad
8/16/2021 1:40:32 PM




#1 Tetrachloro-m-xylene

R.T.: 5.568 min
Delta R.T.: -0.003 min
Response: 17239869
Conc: 16.98 ng/ml




#12 4,4'-DDE

R.T.: 7.637 min
Delta R.T.: -0.008 min
Response: 821182
Conc: 1.70 ng/ml



#12 4,4'-DDE

R.T.: 8.564 min
Delta R.T.: -0.007 min
Response: 1191045
Conc: 1.13 ng/ml




#16 4,4'-DDD

R.T.: 8.222 min
Delta R.T.: -0.008 min
Response: 180954
Conc: 0.42 ng/ml m

Instrument :
ECD_L
ClientSampleId :
TP2-1-6FT


Manual Integrations
APPROVED

mohammad
8/16/2021 1:40:32 PM




#16 4,4'-DDD

R.T.: 9.104 min
Delta R.T.: -0.007 min
Response: 493713
Conc: 0.58 ng/ml m




#17 4,4'-DDT

R.T.: 8.483 min
Delta R.T.: -0.009 min
Response: 841941
Conc: 1.73 ng/ml



#17 4,4'-DDT

R.T.: 9.423 min
Delta R.T.: -0.005 min
Response: 921712
Conc: 0.97 ng/ml




#28 Decachlorobiphenyl

R.T.: 10.404 min
Delta R.T.: -0.010 min
Response: 10263191
Conc: 17.86 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP2-1-6FT

Manual Integrations
APPROVED

mohammad
8/16/2021 1:40:32 PM



#28 Decachlorobiphenyl

R.T.: 11.488 min
Delta R.T.: -0.007 min
Response: 16849536
Conc: 17.37 ng/ml