

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081522\
 Data File : PL077109.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2022 16:20
 Operator : AR\AJ
 Sample : N4200-07
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 06:55:16 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL080822.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 09 09:13:15 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.418	2.635	7215289	19008046	7.358	6.640
28)	SA Decachlor...	8.906	7.739	7112918	20002322	9.422	8.885
Target Compounds							
3)	MA gamma-BHC...	0.000	3.444	0	28171	N.D.	0.007 #
12)	B 4,4'-DDE	0.000	5.064	0	216035	N.D.	0.075 #
14)	MA Endrin	0.000	5.482f	0	253787	N.D.	0.100 #
15)	B Endosulfa...	6.709f	0.000	220742	0	0.257	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081522\
Data File : PL077109.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2022 16:20
Operator : AR\AJ
Sample : N4200-07
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 06:55:16 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL080822.M
Quant Title : GC Extractables
QLast Update : Tue Aug 09 09:13:15 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.418 min
Delta R.T.: -0.002 min
Response: 7215289
Conc: 7.36 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.635 min
Delta R.T.: -0.001 min
Response: 19008046
Conc: 6.64 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 0.000 min
Exp R.T. : 4.220 min
Response: 0
Conc: N.D.



#3 gamma-BHC (Lindane)


R.T.: 3.444 min
Delta R.T.: -0.010 min
Response: 28171
Conc: 0.01 ng/ml



#12 4,4'-DDE


R.T.: 0.000 min
Exp R.T.: 6.080 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#12 4,4'-DDE

R.T.: 5.064 min
Delta R.T.: -0.004 min
Response: 216035
Conc: 0.07 ng/ml




#14 Endrin

R.T.: 0.000 min
Exp R.T.: 6.463 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 5.482 min
Delta R.T.: 0.019 min
Response: 253787
Conc: 0.10 ng/ml



#15 Endosulfan II


R.T.: 6.709 min
Delta R.T.: 0.018 min
Response: 220742
Conc: 0.26 ng/ml

Instrument :
ECD_L
ClientSampleId :




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 5.759 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.906 min
Delta R.T.: 0.000 min
Response: 7112918
Conc: 9.42 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.739 min
Delta R.T.: -0.002 min
Response: 20002322
Conc: 8.89 ng/ml